

## Prevention of Tooth and Gum Damage: Novel Application of Airtraq Laryngoscope

Conventional laryngoscopy with the Macintosh Laryngoscope has been the gold standard for tracheal intubation since 1940s.<sup>1</sup> Failed attempts to intubate the trachea and secure airway, both in the operating room and outside, still remains a cause of anesthesia-related mortality and morbidity.<sup>2-5</sup> Investigation is, therefore, in progress to find the best methods and/or equipment to augment the success rate and lessen the potential complications of tracheal intubation.<sup>6,7</sup>

The Airtraq Laryngoscope (Prodol, Meditec, Viscaza, Spain) (Figure 1) was first reported as a useful device for indirect laryngoscopy and tracheal intubation in 2006.<sup>8</sup> Applications of the Airtraq are not limited to difficult intubation.<sup>9</sup> Some studies have shown its advantages over the Macintosh Laryngoscope with regard to “first attempt success” rate and less intubation time.<sup>10</sup> Others have focused on the Airtraq role as recommended intubation device in special situations such as emergent intubation in the operating room<sup>11</sup> and outside the operating room<sup>12, 13</sup> as well as cervical spine immobilization.<sup>14</sup> These studies have sought to assess the hemodynamic response to tracheal intubation, hypoxemia, esophageal intubation, regurgitation, airway trauma, and cardiac arrest. None of these studies has, however, mentioned the probable role of the Airtraq in decreasing dental trauma.

Tooth damage is one of the most prevalent complications of tracheal intubation and the most common cause of litigations.<sup>15-17</sup> Dental damage occurs as a consequence of several predisposing factors such as preexisting dental pathology; fragile tooth structure; mobility of teeth; presence of prostheses, which increases the likelihood of damage; anterior edentulism, which puts the patient at increased risk of damage to the gums (Figure 2); and the space created after tooth loss, which causes blade lodging between the teeth during conventional laryngoscopy (Figure 3). The conventional direct laryngoscopy exerts more pressure on the maxillary incisors than does indirect videolaryngoscopy.<sup>18</sup> Such problems can be minimized using the Airtraq Laryngoscope for tracheal intubation. The Airtraq blade is not as hard as the metallic blade of the Macintosh Laryngoscope; in addition, its special curve and the view it provides for intubation facilitate intubation without imposing extraordinary pressure on the teeth and soft tissue of the oral cavity (Figure 4).



Figure 1. Airtraq laryngoscope



2A



2B



2C



3A



2D



3B

Figure 3. Macintosh Laryngoscope lodging between the teeth in a patient with anterior edentulism (A) before and (B) after tracheal tube placement



2E

Figure 2. Different features of poor dentition in patients undergoing laryngoscopy and tracheal intubation



4A



4B



4C



4D

Figure 4. Airtraq Laryngoscope prevents trauma to teeth and gum before (A, B&C) and after (D) tracheal tube placement

Poor dentition, the most common cause of dental damage, is more probable in higher age.<sup>19</sup> The recent substantial

increase in the aging population and concomitant rise in the number of candidates for surgical procedures such as coronary artery bypass surgery requires that novel means and ways be devised to reduce the risk of dental damage in laryngoscopy. Further research is needed to shed more light on the pros and cons of the usage of the Airtraq Laryngoscope in patients with malodentition and/or anticipated difficult tracheal intubation.

## References

1. Macintosh RR. A new laryngoscope. *Lancet* 1943;1:205.
2. Cook TM, Woodall N, Frerk C; Fourth National Audit Project. Major complications of airway management in the UK: results of the Fourth National Audit Project of the Royal College of Anaesthetists and the Difficult Airway Society. Part 1: anaesthesia. *Br J Anaesth* 2011;106:617-631.
3. Peterson GN, Domino KB, Caplan RA, Posner KL, Lee LA, Cheney FW. Management of the difficult airway: a closed claims analysis. *Anesthesiology* 2005;103:33-39.
4. Jaber S, Amraoui J, Lefrant JY, Arich C, Cohendy R, Landreau L, Calvet Y, Capdevila X, Mahamat A, Eledjam JJ. Clinical practice and risk factors for immediate complications of endotracheal intubation in the intensive care unit: a prospective, multiple-center study. *Crit Care Med* 2006;34:2355-2361.
5. Griesdale DE, Bosma TL, Kurth T, Isac G, Chittock DR. Complications of endotracheal intubation in the critically ill. *Intensive Care Med* 2008;34:1835-1842.
6. Langeron O, Amour J, Vivien B, Aubrun F. Clinical review: Management of difficult airways. *Crit Care* 2006;10:243.
7. American Society of Anesthesiologists Task Force on Management of the Difficult Airway. Practice guidelines for management of the difficult airway: an updated report by the American Society of Anesthesiologists Task Force on management of the Difficult Airway. *Anesthesiology* 2003;98:1269-1277.
8. Maharaj CH, O'Croinin D, Curley G, Harte BH, Laffey JG. A comparison of tracheal intubation using the Airtraq or the Macintosh laryngoscope in routine airway management: a randomised, controlled clinical trial. *Anaesthesia* 2006;61:1093-1099.
9. Maharaj CH, Costello JF, Harte BH, Laffey JG. Evaluation of the Airtraq and Macintosh laryngoscopes in patients at increased risk for difficult tracheal intubation. *Anaesthesia* 2008;63:182-188.
10. Lu Y, Jiang H, Zhu YS. Airtraq laryngoscope versus conventional Macintosh laryngoscope: a systematic review and meta-analysis. *Anaesthesia* 2011;66:1160-1167.
11. Dhonneur G, Ndoko S, Amathieu R, Housseini LE, Poncet C, Tual L. Tracheal intubation using the Airtraq in morbid obese patients undergoing emergency cesarean delivery. *Anesthesiology* 2007;106:629-630.
12. Woollard M, Lighton D, Mannion W, Watt J, McCrea C, Johns I, Hamilton L, O'Meara P, Cotton C, Smyth M. Airtraq vs standard laryngoscopy by student paramedics and experienced prehospital laryngoscopists managing a model of difficult intubation. *Anaesthesia* 2008;63:26-31.
13. Matic AA. Use of the Airtraq with a fiberoptic bronchoscope in a difficult intubation outside the operating room. *Can J Anaesth* 2008;55:561-562.
14. McElwain J, Laffey JG. Comparison of the C-MAC®, Airtraq®, and Macintosh laryngoscopes in patients undergoing tracheal intubation with cervical spine immobilization. *Br J Anaesth* 2011;107:258-264.
15. Owen H, Waddell Smith I. Dental trauma associated with



- anaesthesia. *Anaesth Intensive Care* 2000;28:133-145.
16. Monaca E, Fock N, Doehn M, Wappler F. The effectiveness of preformed tooth protectors during endotracheal intubation: an upper jaw model. *Anesth Analg* 2007;105:1326-1332.
  17. Gaudio RM, Barbieri S, Feltracco P, Tiano L, Galligioni H, Uberti M, Ori C, Avato FM. Traumatic dental injuries during anaesthesia. Part II: medico-legal evaluation and liability. *Dent Traumatol* 2011;27:40-45.
  18. Lee RA, van Zundert AA, Maassen RL, Willems RJ, Beeke LP, Schaaper JN, van Dobbelen J, Wieringa PA. Forces applied to the maxillary incisors during video-assisted intubation. *Anesth Analg* 2009;108:187-191.
  19. Gaudio RM, Feltracco P, Barbieri S, Tiano L, Alberti M, Delantone M, Ori C, Avato FM. Traumatic dental injuries during anaesthesia: part I: clinical evaluation. *Dent Traumatol* 2010;26:459-465.

**Mahdi Najafi, M. D.**

Assistant Professor of Anesthesiology,  
Tehran University of Medical Sciences,  
Tehran Heart Center,  
North Kargar Street,  
Tehran,  
Iran.  
1411713138.  
Tel: +98 21 88029674.  
Fax: +98 21 88029724.  
E-mail: najafik@sina.tums.ac.ir.

Archive of SID