# Correlation of Central Corneal Thickness with Schiotz and Goldmann Tonometry

## Mohammad-Reza Razeghinejad, MD<sup>1</sup>; Heydar Amini, MD<sup>2</sup>; Fatemeh Torkaman, MD<sup>1</sup>; Hamid-Reza Tabatabaee, MSc<sup>1</sup>

<sup>1</sup>Shiraz Medical University, Shiraz, Iran; <sup>2</sup>Tehran Medical University, Tehran, Iran

**Purpose**: To investigate the correlation between central corneal thickness (CCT) and intraocular pressure (IOP) measured by Schiotz and Goldmann tonometers.

**Methods:** CCT and IOP were determined by ultrasonic pachymetry and with the Goldmann and Schiotz tonometers, respectively. The correlation between IOP measured by Goldmann and Schiotz tonometers on one hand and CCT on the other, was determined by regression analysis.

**Results**: Overall, 168 eyes of 168 glaucoma patients including 85 male (50.6%) and 83 female (49.4%) subjects with mean age of 54.6±19.8 (range 7-85) years were evaluated. Mean CCT was 547.6±53.02 (range: 446-848)  $\mu$ m without any significant difference between men and women (P=0.811). IOP was correlated with CCT with both types of tonometry: every 20  $\mu$ m change in CCT was associated with 1.9±1.4 mmHg and 1.54±0.95 mmHg change in IOP as measured by the Goldmann and Schiotz tonometers respectively with no statistically significant difference between the two devices (P=0.6). **Conclusion:** IOP measurement by Schiotz tonometry is affected by CCT to the same extent as Goldmann applanation tonometry.

Iran J Ophthalmic Res 2007; 2 (2): 101-106.

**Correspondence to:** Mohammad-Reza Razeghinejad, MD. Associate Professor of Ophthalmology; Department of Ophthalmology, Khalili Hospital, Shiraz, Iran; Tel/fax: +98 711 6279373, e-mail: razeghinejad@yahoo.com, mrazeghi@sums.ac.ir

## INTRODUCTION

Glaucoma is a neurodegenerative disorder of the optic nerve and one of the leading causes of blindness worldwide.<sup>1</sup> The most important risk factor for glaucoma is elevated intraocular pressure (IOP) making its accurate measurement essential for disease detection and patient follow-up.<sup>2,3</sup> Although tonometry alone is insufficient for glaucoma screening, it is still useful in case finding.<sup>4</sup> The Goldmann applanation tonometer (GAT) has been universally accepted as the gold standard because of its validity and reproducibility.<sup>5,6</sup> However, the accuracy of GAT depends on central corneal thickness (CCT).<sup>5-7</sup> The importance of CCT and its effect on IOP measurement is going to increase in near future; widespread performance of laser refractive surgery and the resulting decrease in corneal thickness will affect the reliability of IOP measurements.

General practitioners are in a favorable position to detect glaucoma and at present, the Schiotz indentation tonometer is the predominant method of tonometery in general practice.<sup>8</sup> This tonometer is also still used by some ophthalmologists in certain circumstances such as examination of children.<sup>9,10</sup> The aim of this study was to evaluate the effect of CCT on Schiotz tonometry which to the best of our knowledge, has not been reported previously.

## METHODS

Subjects in this prospective study were recruited from the glaucoma service of the ophthalmology department at Shiraz University of Medical Sciences, Shiraz, Iran. The study was approved by the local ethics committee and included 168 glaucoma patients enrolled from April 2005 to January 2006. Although CCT and IOP were measured in both eyes, only data from left eyes were analyzed. Patients were excluded if they had any corneal pathology (corneal scar, edema, keratoconus), previous penetrating keratoplasty or refractive surgery and if suspected of harboring ocular surface infections.

The examinations were performed in the following order: biomicroscopy, Goldmann applanation and Schiotz tonometry followed by central corneal thickness measurement. The subjects were given 10 minutes of rest throughout the series of IOP measurements. Care was taken to keep the patients comfortable and they were encouraged not to squeeze their eyelids or hold their breath. IOP was measured in the right followed by the left eye, with GAT mounted on a slitlamp and then with a Schiotz tonometer using the 5.5 gram weight and if needed the 7.5 and 10 gram weights. Measurements were converted to mmHg based on the 1955 Friedenwald conversion table. One investigator (M.R.) performed all measurements. Before each measurement with the Goldmann tonometer, a nurse adjusted the tonometer gauge to 15 mmHg. The nurse recorded the measured IOP without informing the investigator of the result. Next, the investigator measured IOP by Schiotz tonometry and recorded the scale readings.

All subjects underwent CCT measurement by one experienced technician at least 30 minutes after IOP measurements with an ultrasonic pachymeter (Paxis, Biovision Inc., Clermont-Ferrand, France). Ten measurements were performed in the center of the cornea and the lowest reading was recorded as CCT. The lowest reading is the most likely to reflect a perpendicular placement of the pachymeter probe and therefore reflects the most accurate measurement.<sup>11</sup>

First of all, outlier analysis was conducted on the collected data. Thereafter we used linear mixed model analysis considering age as a covariate to eliminate its confounding effect. Non-linear regression analysis was conducted for modeling the correlation of Goldmann and Schiotz tonometers with CCT. P values <0.05 were considered statistically significant.

## RESULTS

This study was conducted on 168 glaucoma patients including 85 (50.6%) male and 83 (49.4%) female subjects with mean age of 54.6 $\pm$ 19.8 (range 7-85) years. Mean age in male and female subjects was 56.4 $\pm$ 20.2 and 51.7 $\pm$ 19.1 years, respectively (P=0.114). Mean CCT was 547.6 $\pm$ 53.02 (range 446-848) µm overall, and 548.6 $\pm$ 50.04 µm and 546.6 $\pm$ 55.9 µm in male and female subjects respectively (P=0.811). Mean CCT in different types of glaucoma are shown in table 1.

Figure 1 shows that IOP measured by GAT tends to be correlated with CCT and is lower in eyes with thinner corneas and higher in eyes with thicker corneas. Age as a covariate was excluded from the models by mixed model analysis. By performing non-linear regression analysis we found that the equation relating IOP measured by GAT and CCT is: IOP (GAT) =55.82-(0.179×CCT)+(0.0002×CCT<sup>2</sup>), R<sup>2</sup>= 0.262, P<0.001 (R<sup>2</sup> is the variance and CCT is central corneal thickness in microns); standard error (SE) of the slope for IOP against CCT was 0.07.

Figure 2 shows that IOP measured by Schiotz tonometry also tends to vary with CCT and is lower in the eyes with thinner corneas and higher in eyes with thicker corneas. Age as a covariate was excluded from the models by mixed model analysis. Nonlinear regression analysis revealed that the equation correlating IOP measured by Schiotz tonometry and CCT is: IOP (Schiotz) =85.27- $(0.289 \times CCT)$ + $(0.0003 \times CCT^2)$ , R<sup>2</sup>=0.355, P<0.001 (R<sup>2</sup> is the variance and CCT is central corneal thickness in microns); standard error (SE) of the slope was 0.21.

The amount of change in IOP per 20  $\mu$ m change in CCT was 1.9 $\pm$ 1.4 mmHg with Goldmann and 1.54 $\pm$ 0.95 mmHg with Schiotz tonometers (P=0.6).

Diagnosis	Number of	CCT (µm)
	Patients (%)	Mean ±SD (Range)
Primary open angle glaucoma	43 (25.6)	535.6±33.1 (450-539)
Chronic angle closure glaucoma	40 (23.8)	542.7±50.2 (451-685)
Glaucoma suspect	26 (15.5)	547.8±41.9 (455-628)
Pseudoexfoliative glaucoma	22 (13.1)	529.2±37.1 (470-642)
Aphakia secondary to congenital cataract	12 (7.1)	632.6±80.3 (535-848)
Normal tension glaucoma	7 (4.2)	522±10.1 (508-537)
Juvenile open angle glaucoma	6 (3.6)	538.1±17.1 (512-559)
Weill–Marchesani syndrome	5 (3.0)	614±38.7 (554-662)
Pigmentary glaucoma	3 (1.8)	569±21 (552-593)
Aniridia	2 (1.2)	511.5±92.6 (446-577)
Primary congenital glaucoma	2 (1.2)	510±11.3 (502-518)
Total	168 (100)	
SD, standard deviation		

Table 1 Central corneal thickness (CCT) in different types of glaucoma



**Figure 1** Scatter plot depicting the correlation of central corneal thickness and intraocular pressure measured by Goldmann applanation tonometry.

IRANIAN JOURNAL OF OPHTHALMIC RESEARCH 2007; Vol. 2, No. 2



Figure 2 Scatter plot depicting the correlation of central corneal thickness and intraocular pressure measured by Schiotz tonometry.

### DISCUSSION

This study revealed that both Goldmann and Schiotz tonometers tend to overestimate and underestimate IOP in thick and thin corneas, respectively and that CCT affected IOP measurements obtained by Schiotz and GAT to the same amount. To the best of our knowledge, this is the first study demonstrating a correlation between CCT and IOP measurement by Schiotz tonometry, although in a case report by Johnson et al<sup>12</sup> high Schiotz readings were related to a thick cornea. These authors reported a 17-year-old girl who had IOP of 35 and 34 mmHg measured with Perkins and Schiotz tonometers, respectively, without any glaucomatous optic nerve damage and normal visual fields. Medical treatment was unsuccessful in substantially lowering IOP. CCT was 900 um in both eyes. Cannulation of the anterior chamber in the left eye revealed an IOP of 11 mmHg.

The literature suggests IOP errors by GAT ranging from 0.22 to 1.42 mmHg for every

20  $\mu$ m of deviation from a normal CCT of 520  $\mu$ m.<sup>5,13-16</sup> In our study the amount of change of IOP measured by GAT for every 20  $\mu$ m change in CCT was 1.9 mmHg, but there is no data to compare the results of Schiotz tonometry (1.54 mmHg for every 20  $\mu$ m difference in CCT). The difference in the amount obtained for GAT with other studies seems to be related to the inclusion of corneas with CCT greater than 675  $\mu$ m. In all of the above-mentioned studies the maximum CCT was 675  $\mu$ m.

One limitation of the present study is that all participants were glaucomatous; conversely this can be the strength of this study which provided a wide range of IOPs and CCTs. We did not measure corneal astigmatism before measuring IOP and this may have some impact on the results obtained from the Goldmann tonometer. However, the large number of patients sampled in the study is likely to mitigate the potential effect of this variable.

Although IOP measurement is insufficient as a screening tool for glaucoma, it is still useful in glaucoma case finding. With the increasing number of patients with iatrogenically thinned cornea by laser refractive surgery, the measured IOP in these patients is lower than the true value. Most individuals who undergo laser refractive surgery are myopic who are at a 2-6 fold increased risk of developing glaucoma, the risk being proportionate to the degree of myopia. This is of particular concern with higher myopic corrections in which laser refractive surgery is associated with more severe corneal thinning and greater IOP underestimation.<sup>17</sup>

The trend after laser refractive surgery is a mean decrease of 0.63 mmHg in IOP measured by GAT per one diopter of correction.<sup>17</sup> According to the results of our study, IOP will be underestimated using the Schiotz tonometer similar to GAT. This fact may be of concern in the referral of subjects for suspected glaucoma by general practitioners who predominantly use the Schiotz device. On the other hand, IOP is overestimated in the presence of a thick cornea; many individuals with elevated Schiotz readings but no other findings suggestive of glaucoma probably have normal IOP and do not need treatment.

As noted the Schiotz tonometer is used by some ophthalmologists in pediatric patients. Measurement of IOP by this tonometer may be misleading in eyes with thicker or thinner CCT such as patients with previous congenital cataract surgery<sup>18</sup> and congenital glaucoma<sup>19</sup> respectively. In our series, we detected relatively thick corneas in patients with Weill-Marchesani syndrome and aphakic patients following congenital cataract surgery which is compatible with previous reports.<sup>18,20,21</sup> Due to the limited number of patients in other types of glaucoma, it would not be proper to compare our findings with previous reports.

In summary CCT seems to affect IOP measurements by Schiotz tonometry and the correlation is similar to that well known for GAT. This finding may be of particular importance in glaucoma case finding by general practitioners who use the Schiotz tonometer especially in myopic patients with previous laser refractive surgery.

### Acknowledgments

The authors wish to thank Ms Fatemeh Ghalebi from the department of ophthalmology for performing the pachymetries. This study was financially supported by a grant from Shiraz University of Medical Sciences.

#### REFERENCES

- 1. McKinnon SJ. Glaucoma: ocular Alzheimer's disease? *Front Biosci* 2003;1:1140-1156.
- 2. Wilson MR. Epidemiological features of glaucoma. *Int Ophthalmol Clin* 1990;30:153-160.
- 3. Horowitz GS, Byles J, Lee J, D'Este C. Comparison of the Tono-Pen and Goldmann tonometer for measuring intraocularpressure in patients with glaucoma. *Clin Exp Ophthalmol* 2004;32:584-589.
- 4. Sponsel WE. Tonometry in question: can visual
- screening tests play a more decisive role in glaucoma diagnosis and management? *Surv Ophthalmol* 1989;33 Suppl):291-300.
- 5. Lleo A, Marcos A, Calatayud M, Alonso L, Rahhal SM, Sanchis-Gimeno JA. The relationship between central corneal thickness and Goldmann applanation tonometry. *Clin Exp Optom* 2003;86:104-108.
- 6. Gunvant P, Baskaran M, Vijaya L, Joseph IS, Watkins RJ, Nallapothula M, et al. Effect of corneal parameters on measurements using the pulsatile ocular blood flow tonograph and Goldmann applanation tonometer. *Br J Ophthalmol* 2004;88:518-522.
- Kaufmann C, Bachmann LM, Thiel MA. Comparison of dynamic contour tonometry with Goldmann applanation tonometry. *Invest Ophthalmol Vis Sci* 2004;45:3118-3121.
- 8. Christoffersen T, Holtedahl K, Ringberg U, Fors T. Can the Tono-Pen replace the Schiotz tonometer in general practice? *Scand J Prim Health Care* 1998;16:238-241.
- 9. Bordon AF, Katsumi O, Hirose T. Tonometry in pediatric patients: a comparative study among Tono-pen, Perkins, and Schiotz tonometers. *J Pediatr Ophthalmol* Strabismus 1995; 32:373-377.
- Dietlein TS, Jacobi PC, Krieglstein GK. Clinical discrepancy between Schiotz and Perkins tonometry in juvenile glaucoma. *Klin Monatsbl Augenheilkd* 1996;209:299-303.[English Abstract].
- Copt RP, Thomas R, Mermoud A. Corneal thickness in ocular hypertension, primary open-angle glaucoma, and normal tension glaucoma. *Arch Ophthalmol* 1999;117:14-16.
- 12. Johnson M, Kass MA, Moses RA, Grodzki WJ. Increased corneal thickness simulating elevated

intraocular pressure. *Arch Ophthalmol* 1978;96:664-665.

- Bron AM, Creuzot-Garcher C, Goudeau-Boutillon S, d'Athis P. Falsely elevated intraocular pressure due to increased central corneal thickness. *Graefes Arch Clin Exp Ophthalmol* 1999;237:220-224.
- 14. Wolfs RC, Klaver CC, Vingerling JR, Grobbee DE, Hofman A, de Jong PT. Distribution of central corneal thickness and its association with intraocular pressure: the Rotterdam Study. *Am J Ophthalmol* 1997;123:767-772.
- 15. Ehlers N, Bramsen T, Sperling S. Applanation tonometry and central corneal thickness. *Acta Ophthalmol (Copenh)* 1975;53:34-43.
- Singh RP, Goldberg I, Graham SL, Sharma A, Mohsin M. Central corneal thickness, tonometry, and ocular dimensions in glaucoma andocular hypertension. *J Glaucoma* 2001;10:206-210.

- 17. Lee GA, Khaw PT, Ficker LA, Shah P. The corneal thickness and intraocular pressure story: where are we now? *Clin Exp Ophthalmol* 2002;30:334-337.
- Amino K, Miyahara S, Tanihara H. Corneal thickness in eyes following pars plana lensectomy for congenital cataracts. *Jpn J Ophthalmol* 2004;48:169-171.
- 19. Henriques MJ, Vessani RM, Reis FA, de Almeida GV, Betinjane AJ, Susanna R Jr. Corneal thickness in congenital glaucoma. *J Glaucoma* 2004;13:185-188.
- 20. Razeghinejad MR, Safavian H. Central corneal thickness in patients with Weill-Marchesani syndrome. *Am J Ophthalmol* 2006;142:507-508.
- Nilforushan N, Falavarjani KG, Razeghinejad MR, Bakhtiari P. Cataract surgery for congenital cataract: endothelial cell characteristics, corneal thickness, and impact on intraocular pressure. *J AAPOS* 2007;11:159-161.