

Original Research Article

Chemical Methodologies 6(2022) 91-102

Chemical Methodologies

Journal homepage: http://chemmethod.com



New Turbidimetric Method for Determination of Cefotaxime Sodium in Pharmaceutical Drugs Using Continuous Flow Injection Manifold Design with CFTS- Vanadium Oxide Sulfate System

Zahraa Dhahir*[©], Nagham Shakir Turkie

College of Science Department of Chemistry, Baghdad University, Baghda, Iraq

ARTICLE INFO

Article history

ABSTRACT

Submitted: 2021-10-11 Revised: 2021-11-06 Accepted: 2021-11-09 Manuscript ID: CHEMM-2110-1386 Checked for Plagiarism: **Yes** Language Editor: Dr. Behrouz Jamalvandi Editor who approved publication: Dr. Hasan Karimi-Maleh

DOI: 10.22034/chemm.2022.2.2

KEYWORDS

Cefotaxime sodium Flow injection Scatter light ISNAG-Fluorimeter instrument

The determination of Cefotaxime sodium pure from and pharmaceuticals (tablets) by continuous flow injection scatter light has been created using a new, simple, rapid, and sensitive approach. Using a handmade ISNAG-Fluorimeter, the technique was predicated on the interaction of Cefotaxime sodium with vanadium oxide sulfate to produce a precipitate. The optimal value had a linear dynamic range of 0.01-20 mmoI/L for the instrument response to Cefotaxime sodium concentration, while the L.O.D was 0.334 g /sample from the stepwise dilution for the lowest concentration in the calibration graph's scatter plot range. For the scatter plot range and the linear dynamic range, the correlation coefficient (r) was 0.9769 and 0.9978, respectively, while percentage linearity (R²% percent) was 95.43 percent and 99.56 percent, respectively. With a concentration of 2.5 mmol/L, the RSD percent for repeatability (n=8) was lower than 0.23% percent for the determination of cefotaxime sodium. The proposed method was effectively used to determine the concentration of cefotaxime sodium in pharmaceutical tablets. The newly developed technique was compared with the conventional methods of analysis, i.e., Turbidemtric method and UV-spectrophotometer method, utilizing the usual addition method and paired t-test. The computed t-value at 95 percent confidence interval from the established approach revealed that there was no significant difference between the quoted figures of any individual firm.



* Corresponding author: Zahraa Dhahir E-mail: <u>zahraa.dhahir1205@sc.uobaghdad.edu.iq</u> © 2022 by SPC (Sami Publishing Company)

Introduction

Cefotaxime sodium is a sodium antibiotic (6R,7R) -3- [(acetyloxy)methyl] ((acetyloxy)methyl) ((acetyloxy)methyl) ((-7- (2Z) 2-(2aminothiazol-4-yl)-2-(methoxyimino) acetyl] amino]-2-(2- aminothiazol-4-yl)-2-(methoxyimino) acetyl] amino] thia-oxo-5-thiaoxo-5-thia-oxo-5-thia-o-1-azabicyclo $\left[4.2.0\right]$ $C_{16}H_{16}N_5NaO_7S_2$, oct-2-ene-2-carboxylate]. It has a molecular weight of 477.447 and is a white or slightly yellow powder that is readily soluble in water but sparingly soluble in methanol. Figure 1 shows the chemical structure of CFTS [1].



Figure 1: chemical structure of CFTS

Cefotaxime is a broad-spectrum antibiotic that is typically used to treat bacterial infections of the skin, soft tissues, and urinary tract. It is a kind of antibiotic that is widely used. Cefotaxime sodium is a cephalosporin antibiotic that belongs to the third generation of cephalosporins. In addition to its antimicrobial action, third-generation Cephalosporins have a broad antibacterial range and resistance to beta-lactamase-producing organisms in addition to its antimicrobial activity, i.e., streptococci, staphylococci, pneumococci, etc. [2]. Cefotaxime failed to permeate the central nervous system and was unsuccessful in treating meningitis, but CFTS penetrates the central nervous system and achieves therapeutic concentrations, which are sufficient for treating meningitis caused by aerobic gram-negative bacteria [3]. It works against both Gram positive and Gram negative bacteria and has a broad spectrum of action. It is generally thought to be safe and effective in the same way as ceftriaxone is. These features are clinically and analytically significant. CFTS may be analyzed using a variety of analytical techniques that can be found in the HPLC-UV literature. detection, spectrophotometric Fluorescence, Electrochemical (Electrochemical-ISE, Electrochemical-MIP, turbidity using Ayah 4Sw-3D-T180o-2N90o-Solar-CFI analyzer [4], fluorescence resonance energy transducer, and continuous flow injection analysis via ISNAGfluorimeter are some of these methods in their various forms and preparations [5]. In this research that was carried out with flow injection analysis with scattering light [6], measuring

diverged beam of light at 0-900 angle was detected by homemade ISNAG- fluorimeter [7-10] via low-pressure mercury lamp as a source and using 2 sides [4 x 2.5cm] solar cells for the determination of CFTS with Vanadium oxide sulfate as a precipitating agent in the presence of Hydrochloride acid solutions.

Apparatus

A homemade ISANG fluorimeter was used with multichannel more than one-line feed. In this part of the research work only two lines, and fourchannels peristaltic pump (Ismatec, Switzerland) and Six-port medium pressure injection valve were used. The output of measurement *i.e*; \bar{Y}_{zi} (mV)-t_{min}(d_{mm}) was plotted by potentiometric recorder to determine the output signals (Siemens, Germany (1- 5 V). HANNA instrument for turbidity measurement was used similar to classical methods.

Materials and methods

Reagent and chemical

All chemicals were used of analytical-reagent and distilled water was used to prepare all the solutions. A standard solution of cefotaxime sodium (CFTS) ($C_{16}H_{16}N_5NaO_7S_2$, molecular weight 477.447, 0.1mol.L⁻¹) was prepared by dissolving 0.477447 g in 100 ml distilled water. We used a stock solution Vanadium oxide sulfat VOSO₄, molecular weight 163.01 g/mol, 8.1505 g g/250 mL, Na₂CO₃ 5.299 g /100 mL, CH₃COONH₄ 3.884 g /100 mL, KCl 7.350 g/100 mL, NaCl 7.313 g/250 mL, NH₄Cl 2.675 g/100 mL, CH₃COOH

57.47 ml/L, HCl 88.28 ml/L and $\rm H_2SO_4$ 55.52 ml/L.

Sample preparation of Cefotaxime sodium (CFTS)

A batch of Twenty tablets were weighed then crushed and grinded; each of the drug containing 1000,500 mg of Cefotaxime sodium was weighed 0.48460 g, 0.40668 g and 0.40688 g (equivalent to 0.477447 g of active ingredient, 20 mmol.L⁻¹) for Cefotaxime-SANAVITA-Germany, Cefotaxime-LDP-Spain & Cefotopox-Russul-India respectively. The powder was dissolved in distilled water. The solution was filtered to remove any undissolved residue that affected the response and complete the volume to 100 mL with distilled water.

Methodology

A two-line manifold reaction design was used and connected with ISNAG-fluorimeter instrument to

estimate the cefotaxime sodium via its reaction with VOSO₄ to form white precipitate, as shown in Figure 2. First line as a carrier stream of distilled water carried the sample segment (140 µL) of 8 mmol/L of Cefotaxime Sodium concentration at 1.8 ml/min flow rate then combined with the second line (VOSO₄ Line) at 1.8 ml/min flow rate via Y-junction point that carried the reagent of VOSO₄ at 7 mmol/L. So, when the reagent (precipitating agent) joined with the incoming drug causing the formation of small crystalline precipitate particulate, followed by the measurements of divergent light via irradiating the precipitated particulate by two wavelengths at 184 nm & 253.7 nm (Prominent) through homemade ISNAG - Fluorimeter instrument (Figure 2). The proposed probable reaction pattern of ion pair is shown in scheme 1 [11,12].



Figure 2: Flow diagram of used manifold throughout this part of research work



Scheme 1: Proposed reaction between CFTS with VOSO₄ to form ion pair

^{93 |} Page

Optimization of reaction pattern parameters

Chemicals parameters, mainly concentration of reagent and type of carrier stream for the reaction of CFTS with VOSO₄ system, as well as physical parameters including sample volume, flow rate and reaction coil were studied using two-line manifold system.

Chemical variables

Effect of vanadium oxide sulphate

Variable concentrations of vanadium oxide sulphate reagent ranging 1-10 mmol/L were prepared at 140 μ L sample volume, which were injected on a carrier stream (distill water). 8 mmol/L concentration of CFTS was used with 1.8 ml/min flow rate for each line. Figure 3.A shows a kind of response versus time of profile (Y_{Zi}(mV)t_{min} (d_{mm}) 1(10)). The increase of VOSO₄ led to an increase in the peak height reaching the value of 5 mmol/L (1552 mV). This increase might be attributed to the increase of small nuclei density which in turn start to grow to form granules as the reflecting light surface toward the detector in a more intense and shorter frequency (blue shift) and falls within solar cell limit of sensing while more than 5 mmol/L causes low sensitive responses due to the possible formation of vacancies or pockets to accommodate unwanted particulate *i.e.*, impurities relative to the nature of precipitate causing spreading or dispersed particulate leading to decrease of diverged light toward the solar cell. This was noticed using slope-intercept method (Figure 3.B); which shows that the selected segment giving the highest value is $a_3 - a_5$ in which 5 mmol/L fell within its boundaries. Also, any chosen concentration of VOSO₄ within this segment can be used as an alternative to use in newly developed methodology of optimization.



Figure 3: Variation of VOSO₄ concentration on A: Profile of CFTS – VOSO₄ system

B: Height of $\bar{Y}_{Zi}(mV)$ energy transducer response in mV with three data points as one segment their interaction and choice

Type of media effect (salt and acid)

Cefotaxime Sodium (8 mmol/L) reaction with VOSO₄ (5mmol/L) was studied in different media (CH₃COONH₄, NaCl, Na₂SO₄, KBr, NaNO₃, NH₄Cl, KCl, HCl, CH₃COOH, H₂SO₄, HNO₃, Tartaric acid (C₄H₆O₆) and Ascorbic acid (C₆H₈O₆)) at 50 mmol/L concentration in addition to aqueous medium to be used as a carrier stream at 1.8

ml/min flow rate to improve the light scattering on precipitate particulate surface. It was found that an increase in S/N- response using H_2SO_4 compared with the use of distilled water and salts medium as a carrier stream, which might be due to the increase of particles size that will act as a reflecting surface at 0-9-° and increase of the

scatter of light. Therefore, H_2SO_4 was chosen as the best carrier stream for the next studies.

Effect of sulfuric acid concentration

Using cefotaxime sodium (8 mmol/L)-VOSO₄ (5 mmol/L) system, we prepared 140 μ L sample volume at 1.8 ml/min flow rate for both lines, and a series of sulfuric acid solution (10- 100 mmol/L) as carrier stream (Figure 4 and Table1). It was found during working with acids (sulfuric acid) an increase of sulfuric acid concentration more than 10 mmol/L caused a decrease in peak height of response. This could be attributed to the penetration of sulfuric acid to the precipitate depth during its crystal growth forming a soft particulate, which caused a decrease the light scattering and its deviation toward the sensor

(detector in ISNAG-Fluorimeter instrument). Therefore, 10 mmol/L of sulfuric acid was chosen as optimum carrier stream for this research. When applying slope-intercept, in spite of the values of a= 1826.27 mV which represented the lowest value, it was chosen as the most appropriate that would represent the first segment for the range 10-50 mmol/L. Also, it was noticed that the excess of acid added would dissolve the precipitate completely or in step wise by increasing the acid concentration. Accordingly, higher values of intercept which represented the segment 2 and segment 3 were neglected. So, the first segment was chosen, having the lowest value for acid concentration.



Figure 4: Effect of variable concentration of H₂SO₄ solution on: Output of (S/N) energy transducer response in mV with chosen segment

Table 1: Data set point obtained for the variation of Ascorbic acid concentration in the determination of CFTS
using CEFTS (8 mmol/L) – VOSO4 (5 mmol/L) –H2SO4 system

[H2SO4] mmol/L	$\bar{Y}_{zi}(mV)$ average of response (n=3)	RSD %	Confidence interval at 95 % $\bar{Y}_{zi}(mV) \pm t_{0.05/2,n-1} \sigma_{n-1} / \sqrt{n}$
10	1800	0.11	1800 ±4.7699
30	1680	0.12	1680±4.7700
50	1640	0.12	1640±4.7699
70	940	0.21	940±5.0184
100	720	0.32	720±5.7637

t_{0.05/2,2}=4.303, \overline{Y}_{Zi} (mV): (S/N) energy transducer response

Physical variables

Variation effect of flow rate

Fixing cefotaxime sodium (8 mmol/L)-VOSO₄ (5 mmol/L)-H₂SO₄ (10mmol/L) system as optimum parameters and variable flow rate from (0.7 - 2.8 ml/min) for each line (carrier stream line and

reagent line) was used for the study. It was noticed that at slow flow rate (0.7 ml/min), its presence in the measuring cell, which in turn caused a display at the base (Δ tB) and apex of the profile, led to a big particle, considered too heavy to deal with the movement of carrier stream through flow cell, causing a delay and elongation

95 | P a g e



Dhahir Z.,/ Chem. Methodol. 2022, X(X), XX-XX

(Figure 5). Next, increasing flow rate reaching to 1.8 ml/min led to an increase in height of responses were obtained, that might be attributed to the increase of diverging of light directed towards the detector while more than 1.8 ml/min caused a decrease in the peak height of responses, most probably due to unavailability of sufficient time for crystal to grow to higher size

particulate that would work as a reflecting mirror for the incident light. Using slope -intercept method showed that the selected segment giving highest value was (a4 - a6) in which 1.8 ml/min fell within its boundaries. Also, any chosen flow rate within this segment could be used as an alternative to use in newly developed methodology of optimization



Figure 5: Effect of flow rate on :A-Response profile $\bar{Y}_{Zi}(mV)$

Variation effect of sample volume

Variable volume of sample loop ranging 40 μ L to 175 μ L by changing the length of the sample loop within the injection value while keeping all other variables are constant (*i.e.* CFTS 8 mmol/L)-VOSO₄ (5 mmol/L)-H₂SO₄ (10 mmol/L) system and 1.8 ml/min flow rate. Table 2 shows the increase of sample volume to 175 μ L leading to increase of peak height and giving a regular

profile of the diverged incident of light while more than 175 μ L led to a decrease of peak height with a wider Δ t of response. So, the sensitivity of measurements was small at more than 175 μ L sample volume. This result was proven through slope- intercept method in which the segment no.2 is the most suitable, that could be used to obtain sufficient acceptable result compromised with the other variables that were chosen.

 Table 2: Set of data for variation of sample segment and related of measurement output using CFTS (8 mmol/L)

 - VOSO4 (4 mmol/L) -H2SO4 (10 mmol/L) system

Length of sample segment cm r=0.5 mm	Sample volume μL	$ar{Y}_{zi}(mV)$ average of response (n=3)	RSD %	Confidence interval at 95 % \bar{Y}_{zi} (mV) ± t0.05/2,n-1 σ_{n-1} $/\sqrt{n}$	Base width Δt _B sec	t* sec	V _{adds} at flow rat cell ml	Concentration mmol/L At flow cell	D.F
5.09	40	900	0.147	900 ± 3.2793	18	8.0	1.12	0.28	28
11.56	90	1320	0.158	1320 ± 5.1923	21	9.0	1.35	0.53	12
14.0	110	1420	0.149	1420 ± 5.2668	24	10.8	1.55	0.57	14.09
18.0	140	1800	0.128	1800 ± 5.7388	24	12.0	1.58	0.71	11.29
19.11	150	1920	0.127	1920 ± 6.0866	27	15.0	1.77	0.68	11.80
22.30	175	2040	0.143	2040 ±7.2543	33	18.0	2.16	0.65	12.34
25.48	200	1640	0.141	1640±5.7637	42	21	2.72	0.59	13.6

96 | P a g e

Dhahir Z.,/ Chem. Methodol. 2022, 6(2) 91-102

Effect of delay reaction coil on S/N response profile Using the other optimum chemical and physical parameters, variable reaction coils from 0 (without)-236 μ L connected after Y- junction point directly in manifold design system were studied. It was noticed using coil at volume 78.5 μ L led to an increase in sensitivity and peak height of response up to 78.5 μ L, while at more than 78.5 μ L it led to a decrease of diverged light with increase of base width and departure time of sample segment from injection valve reaching the measuring cell. This might be attributed to an increase of diffusion and dispersion of precipitate particulate segment causing loss of some of the reflecting surface. Therefore, delayed reaction coil at volume 78.5 μ L was used in this research. This result corresponds with slope – intercept method which shows that the selected segment giving the highest value of intercept is a₂-a₅ (*i.e.*, segment no.2), in which any point within it can be used to obtain acceptable results and high sensitivity (figure 6).



Figure 6: Variation effect of Coil volume on: Height of \bar{Y}_{Zi} (mV): (S/N) energy transducer responV, three segments (Three-point data) and chosen segments

Studying the variation of energy transducer output of diverged light versus cefotaxime sodium concentration

Using fixed parameters optimization achieved at previous section, we prepared a set of CFTS solutions (0.01-20 mmol/L) for CFTS-VOSO₄ (5 mmol/L)-H₂SO₄ (10 mmol/L) system, in which the direct proportion up to 20 mmol/L between the variation of concentration of precipitate particulate formation and output of diverged light was expressed as an $\overline{\mathbf{Y}}_{Zi}$ (mV). Many variables, such as refraction, absorbance, reflection, and divergent light from inside the precipitated particles, might have increased as the beam of light diffused inside the particles. All values were measured at an angle 0-90° and represented by scatter plot at range 0.01-20 mmol/L, in which we gave a coefficient correlation of $\overline{\mathbf{Y}}_{Zi}$ (mV) versus t_{min} (d_{mm})1(10) of 0.9769 with coefficient of determination of 0.9543 and linear dynamic range 0.01-11.5 mmol/L. A short range that should be used to improve the assessment of mathematical formulation and the best Linear with correlation equation the between cefotaxime sodium concentration against

diverged light as a dependent variable with r=0.9978 and % capital R-squared of 99.56% (table 3), and the linear equation of the form. \hat{Y}_{Zi} =29.827±59.417+256.757±8.713 [CFTS] mmol/L was able to explain much of obtained results by 99.56% from n= 19 as the outcome of linear dynamic. The assessment evaluation of the new developed methodology for the determination of cefotaxime sodium was compared with available references of the two methods including UV-Spectrophotometric method and turbidimetric method based on:

1- UV. Spectrophotometric method: using a quartz cell to detect absorbance for the range of correlation concentration 0.001-0.8 mmol/L at max = 260 nm [0.15 mmol/L]. With a correlation value of 0.9983 and a percent capital square of 99.66 percent, the scatter plot and best linear range ranged from 0.001-0.5 mmol/L, n=15 (no.of measurements). The variable data treatments are shown in Table 3;

2- turbidometry method; which is based on the reaction between $VOSO_4$ as a precipitating agent with optimum concentration 6 mmol/L in aqueous medium and the drug of cefotaxime sodium for a variable range of concentration 0.1-15mmol/L. The results are tabulated in Table 3 at

97 | Page

Dhahir Z.,/ Chem. Methodol. 2022, 6(2) 91-102

confidence level 95 % using first degree equation of the form of:

 $\hat{Y}_i(NTU) = a + b [CFTS] mmol/L$

The best linear ranges extend from 0.1-10 mmol/L of n= 13 with correlation coefficient of 0.9991 and % capital R-squared = 0.9982 % (Table 3).

Intercept Slope

Table 3: Different ranges using linear regression for the variation of S/N energy transducer response and usingspectrophotometer and Turbidimeter method (classical methods) with Cefotaxime sodium concentration usingfirst degree equation of linear $\hat{Y}=a+bx$ at optimum condition for both methods

Type of mode	Range of [CFTS] mmol/L (n)	\hat{Y}_i =a±s _a t+b±s _b t[CFTS] mmol/L at confidence interval 95 %, n-2	r, r ² , R ^{2 %}	t _{tab} at 95 % ,n-2	Calculated t-value $\frac{r}{\sqrt{n-2}}$ $\sqrt{1-r^2}$	
		Developed method using ISNAG-fluorimete	er			
		UV- Spectrophotometer at λ_{max} = 260nm				
	Γ	Turbidometry method (NTU) Using Hanna Inst	rument			
Linear range or Liner dynami c range	0.01-11.5 (19)	29.827±59.417+256.757±8.713 [CFTS] mmol/L	0.9978, 0.9956, 99.56	2.110 < 62.176		
	0.001-0.5(15)	0.041±0.035+3.904±0.136 [CFTS] mmol/L	0.9983, 0.9966, 99.66	2.160< 62.060		
	0.1-10(13)	39.254±11.994+84.389±2.356 [CFTS] mmol/L	0.9991, 0.9982, 99.82	2.201< 78.826		
Scatter plot	0.01-20(22)	0.9769, 0.9543, 95.43	2.08	36 < 20.430		
	0.001-0.8(18) 0.185±0.168+2.894±0.455 [CFTS] mmol/L		0.9588, 0.9193, 91.93	2.120 < 13.499		
	0.1-15(15)	15(15) 82.116±60.887+68.805±8.723[CFTS] mmol/L		2.160 < 17.038		

 \hat{Y}_i : Estimated response (n=3) in mV for newly and developed without unite for spectrophotometric method or in NTU for turbidimetric method (classical method) for n=3 expressed as an average peaks heights or turbidimetric value of linear equation of the form $\hat{\mathbf{Y}} = \mathbf{a} + \mathbf{b} \mathbf{x}$. Correlation coefficient. r: r²:Coefficient of determination, R²:% capital R-R²= explain variation souared. / total variation,[CFTS] concentration of Cefotaxime sodium; Developed method: using ISNAGfluorimeter, n:no.of measurements, t_{tab.}=t 0.05/2,n-2,

volume of measuring cell 1 ml for UV-Sp. and 10 ml for turbidimetric.

Limit of detection (LOD)

The Limit of detection of cefotaxime sodium is calculated by three different approaches depending on the practical research, *i.e.* developed method using VOSO₄ (5 mmol/L) - H_2SO_4 (10 mmol/L) system, 175 µL sample volume and two classical methods (UV-Spectrophotometer & Turbidimetry) (Table 4).

Table 4: Detection Limit of CFTS using 175 μL as an injection sample and optimum parameters using CFTS-VOSO₄ (5 mmol/L)-H₂SO₄ (10 mmol/L) system of developed method and two classical methods



Dhahir Z.,/ Chem. M	Methodol. 2022	, 6(2) 91-102
---------------------	----------------	---------------

Practically b minimu Newly developed method (0.004) mmo/L	ased on the gradual dil m concentration in sca UV- Spectrophotometer (0.0005) mmol/L	ution for the tter plot Turbidimetry method (0.05) mmol/L	Theoretical based on the value of slope x=3S _B /slope`	Theoretical based on the linear equation $\hat{Y} = Y_b + 3Sb$	Limit of quantitative L. O. Q Ŷ=Y _b +10S _b	
0.3342	0.2387	238.723/	0.7322	68.687	228.97	
µg/sample	µg/ sample	sample	μg/ sample	μg/ sample	µg/ sample	

 \hat{Y} : Estimated response (mV), X value of LOD based on the slope (depending on analytical range), S_B standard deviation of blank (n=13), Y_b: average response for blank= intercept (a), S_b: standard deviation equal to S_{y/x} (residual), (LOD depending on linear equation of linear range due to low S_{y/x}) and volume of measuring cell 1 ml for UV-Sp. and 10 ml for turbidimetric.

Repeatability

The reproducibility of the determination of cefotaxime sodium was investigated using an ISNAG-fluorometer to detect the divergence of incoming light, which was generated by the reaction of cefotaxime sodium with VOSO₄ in the presence of H_2SO_4 . The relative standard deviation is expressed as a percentage (RSD

percent 0.23 percent), which is proportional to the measurement's repeatability. In optimal parameters, eight consecutive injections of CFTS at fixed concentrations (2 and 5 mmol/L) were assessed. All of the information gathered is summarized in Table 5.

[CFTS] mmol/L	Υ _{zi} (mV) average of responses (n=8)	RSD %	Confidence interval at 95% Ῡ _{zi} (mV)± t _{0.05/2/n-1} σ _{n-1} / √n
2	456	0.221	456± 0.8445
5	1496	0.089	1496±1.1204

 \bar{Y}_{Zi} (mV): Energy transducer response expressed, t_{0.05/2,7}=2.365, n= number of injection

Determination of Cefotaxime sodium in drugs using a homemade of ISNAG-fluorometer analyzer Cefotaxime sodium was determined in three distinct samples of medicines from three different firms using the newly developed technology (ISNAG-fluorometer). (Cefotaxime-1000mg-SANAVITA-Germany, CEFOTAXIME -500 mg-LDP-Spain and CEFOTOPOX- 500mg-RussuL-India). The continuous flow injection was analyzed by homemade ISNAG-fluorometer analyzer, in which mercury tube lamp was as a source and four solar cells at each two side (0-90°) as a detector; this was compared with two classical methods including UV-Spectrophotometric via the measurements of absorbance at λ_{max} = 260 nm and turbidometrymethod via turbidity-meter- HANNA-Taiuan. Each method depends on the preparation of each drug: 20 mmol/L equivalent to 0.4774 g of active ingredient in 50 ml by transferring 1 ml to each of five volumetric flasks (25 ml), followed by addition of gradual volumes of standard solution (50 mmol/L) of Cefotaxime sodium (0, 1, 1.5, 2 and 2.5 ml) to obtain 0, 2, 3, 4 and 5 mmol/L concentration.

Table 6. A and B show the results mathematically treated [3-5], using two or three different methods, *i.e.* developed method, UV-Spectrophotometric method and turbidometry method, and with a realistic active component content at a 95% confidence level and determination efficiency in addition to t-test.

First test: We used individual t-test to compare the means of practicality of active ingredient (using newly developed methodology) with claimed value (*i.e.*, μ = 500 mg or 1000 mg) as shown in Table 4.B (Column 5)

Null hypothesis: There is no significant different between the means (\overline{w}_i (SANAVITA, 1.0324g, Germany), (LDP, 503.265 mg, Spain), (RussuL,

Dhahir Z.,/ Chem. Methodol. 2022, 6(2) 91-102

488.998 mg, India) (average weight)) obtained from three source of three different companies and claim value [15, 16] (μ =500mg or 1000 mg).

There is a substantial discrepancy between the means and claim value in the Alternative hypothesis.

From the results obtained, it seems that there was no significant difference between measured

mean values (\overline{W} i) compared with official Quoted value ($\mu_{500 \text{ mg or } 1000 \text{ mg}}$) except sample no.3 possibly due to the interferences that cause a random error that exist with this sample (*i.e.*, sample no.3).

Table 6. A: Standard addition results for the determination of Cefotaxime sodium in three samples of drug using

 ISNAG-fluorimeter analyzer for developed method and two classical methods

		Type of method											
			Dev	eloped method using	ISN	AG F	luo	rim	eter	(mV)			
		UV	- Spectrophotometry	y Classical method Ab	sorł	oanc	ce m	eas	urer	nent at $\lambda \max = 260 \text{ nm}$			
		Turbidometry method (Classical method) NTU											
e		Confidence											
ldm		interval											
sai	Commercial Name, Company Content Country	For the	Sample	Theoretical	Гheoretical				Equation of standard				
o. of		average	equivalent to	content for the						addition at 95% for n-2	r		
Ň		Weight of	0.4774 gm	active ingredient							r2		
		Tablet	(20 mmol/L) of	at 95%		1		1			1		
		$WI \pm 1.96 \sigma_{n}$	ingradiant	(mg) Wi+1.96 $\sigma = 1/2/n$	0	-	1.5		2.5	$\hat{Y}_{Zi}=a\pm s_at+b\pm s_bt$			
		1/ VII at 95%	$W_{i}(\sigma)$	WIE1.90 On-1 / VII		01		2		[CFTS]mmol/L	5.00/		
		(g)	WI (6)		0	2	3	4	5		R ² %		
		(8)								213 027+92 747+257 91	0.9982		
		taxime ANITA TS= 1.0150±0.005 00 mg 8 rmany			23	89	10:	126	149	9±28.222	0.9965		
	Cefotaxime SANANITA CFTS= 1000 mg Germany		2				0	0	0	60	96	[CFTS]mmol/L	99.65
										0.237±0.035+0.299±0.0	0.9998		
			5 0.4846	1000±5.7143	0.23	0.84	1.1	1.4	1.7	11 [CFTS]mmol/L	0.9996		
1							б	2	ω		99.96		
1					70	240		415		68.405±10.885+85.784±	0.9998		
							320 240		498	3.312 [CFTS] mmol/L	0.9996		
											99.96		
		EFOTAXIME LDP CFTS= 6					~		~	85.810±7.433+106.568±	0.9999		
					78	296	408	510	520	2.262 [CFTS]	0.9999		
	CEFOTAXIME		KIME									99.99	
	LDP				0	0	-	ь	<u>ц</u>	0.233±0.048+0.280±0.0	0.9996		
	CFTS=		0.4067	500±10.0963	.22	.81	.09	.34	.63	14 [CFTS]mmol/L	0.9992		
2	500 mg										99.92		
	Aspain				8	2	ω	4!	σ	79.054±10.908+95.338±	0.9998		
					ö	70	55	5	50	2.219 [CF12]IIIII0I/L	0.9996		
										147 027+31 65+187 010	0.9996		
					10	ы	70	91	10	+9 630 [CETS]mmo]/I	0.9992		
	CEFOTOPOX				сi	ω	0	0	88		99.92		
	RussuL									0 209+0 082+0 263+0 0	0.9987		
	CFTS=	0.4261±0.011	0.4069	500±13.3771	0.2	0.7	0.9	1.2	1.0	25 [CFTS]mmol/L	0.9973		
_	500mg	4			2	74	71 8(ω	6	- []/	99.73		
3	India					1					0.9998		
					60	22(30(378	45(62.216±9.281+78.351±2	0.9996		
					Ľ	0		8	0	824 [CF15]mmol/L	99.96		

 \tilde{Y}_i : Energy transducer in mV for developed method, without unite for UV-Spectrophotometric method and Turbidimetric method in NTU, for (n=3), r: correlation coefficient, r² coefficient of determination, R²%: Percentage capital R- squared: R² explain variation as a percentage / total variation, t_{0.05/2, ∞} = 1.96 at 95 %, t_{0.05/2, 3} = 3.182 for n=5

100 | P a g e

Dhahir Z.,/ Chem. Methodol. 2022, 6(2) 91-102

Second test: We used paired t-test at α = 0.05 of three drugs from different manufacturing. The comparison was made between developed method using ISNAG-fluorimeter analyzer (depending on scattered light ± 90°) and turbidimetry via turbidity-meter, HANA, (Taiuan) as well as UV-spectrophotometer at λ_{max} = 260 nm. The results of comparison are summed up in Table 6. B (Column 6).

The Assumption as follows [13,14]: Null hypothesis:

There is no significant difference between the means of turbidity method and ISNAG-fluorimeter analyzer

Alternative hypothesis:

 $H_1: \mu_{ISNAG-fluorimeter} \neq \mu_{UV-spectrophotometry} \neq \mu_{turbidimetry}$

As shown in Table 4.B (column 6) due to tcal less than t_{tab} (4.303), there were no significant differences between the developed method and two classical methods (UV-spectrophotometric and turbidimetric methods) for determining cefotaxime sodium in pharmaceutical drugs at 0.05 (95% percent confidence level).

H₀: $\mu_{\text{ISNAG-fluorimeter}} = \mu_{\text{UV-spectrophotometry}} = \mu_{\text{turbidimetry}}$

Table 6. B: Summary of results for practical content, (Rec %) efficiency for determination of Cefotaxime sodiumin three samples of drugs and t-test for comparison between two methods using t-test includes (Individual t-test& paired t-test)

		Type of method							
	De	Developed method using ISNAG Fluorimeter (mV)							
	UV- Spectrophotome	tric Classical method Absorbanc							
]	Surbidometry method (Classical							
e	Practical	Weight of CFTS in each			Pai	red t –test			
ldu	concentration	sample			Compared between two				
sar	(mmol.L-1)	(g)		Individual t-test	me	thods			
of	In 25 ml	$\mathbf{W}_{i(g) + 4.303 \text{ g n} - 1} / \sqrt{\mathbf{n}}$		For compared					
No.	concentration	Woight of CETS	Efficiency of	between claim value					
	(mmol.L-1)	ol L-1) in tablet		& practical value					
	in 50 ml	Ŵ	Rec.%	$(\mathbf{\bar{W}}_{i(ma)}) \sqrt{n}$	tcal=	t _{tab} at			
	Practical weight of	i(g or mg) ±4.303σn-		$(1(112)-\mu)^{-1}/0$	$\overline{\mathbf{w}} \sqrt{\mathbf{n}}$	95%confid			
	CFTS in	$\sqrt{\mathbf{n}}$		11 1	d V-	ence level			
	(g)	1/ -			/σ*n-1	(n-1)	_		
	0.8259	0.4929±0.0390	102.24						
	0.4929	1.03245 g ±0.0820	103.24						
	0.7929	0.4732+0.0043							
	19.8240		99.12	1.70 << 4.303	Paired t-test for development with UV- Sp.				
1	0.4732	$0.99119 \text{ g} \pm 0.0091$							
	19.9350	0.4759±0.0043	99.68						
	0.4759	0. 9968 g ± 0.0091							
	0.8052	0.4805 ± 0.0057	100 (5		α= σ	- 8.2010			
	0.4805	503.2650 mg ± 5.9870	100.65		n-1	= 8.4720			
	0.83220	0.4967 ± 0.0051			/-1.0090	$< t_{tab}$			
2	20.8050	520.1310 mg ±5.2900	104.03	2.347 << 4.303			-		
	0.8290	0.4949±0.0031							
	20.7290	518.2490 mg ± 3.2800	103.65		Datard	+ + + C			
	0.7824	0.4669±0.0098			developme	nt with turb.			
	19.5599	488.008 mg + 10.2820	97.79		X				
	0.4669	400.990 Ilig ±10.2020			d=	-7.4141			
	0.7951 19.8780	0.4745±0.0020	99 39	/-4 604/>> 4 303	σ_{n-1}	= 7.4752			
3	0.4745	496.9580 mg ±2.1200	<i></i>	/ 1.001/ 22 1.303	/-1.7179	9/ < 4.303			
	0.7941	0.4739±0.0030			t _{cal}	< t _{tab}			
	19.8520	496 2920 mg + 3 1220	99.26						
	0.4739	170.2720 mg ± 5.1220					1		

 μ : claim value =1000, 500mg, \overline{W} i:mean of practical weight (n=3), \overline{W} d: average of different between two methods (developed method &*Turbidimetric (NTU)Using Hanna Instrument or UV-Spectrophotometry using Shimadzu (Japan) (classical methods), for n(No. of samples) =3 , σ_{n-1} : standard deviation, σ^*_{n-1} :standard deviation of difference (paired t-test), t_{0.025,2}= 4.303.

Dhahir Z.,/ Chem. Methodol. 2022, 6(2) 91-102

Conclusion

A new turbidimetric, simple, sensitive, accurate and fast method for determination of cefotaxime sodium in pharmaceutical drugs (tablets) was developed using a newly developed homemade ISNAG-fluorimeter -CFIA. The comparison between this works with classical turbidometry method via the t-test (as the comparison tools) showed that with no doubt the newly developed method (ISNAG-fluorimeter procedure) was as good as the classical method.

Acknowledgments

I would like to express my deepest gratitude to Prof. Dr. Issam M.A. Shakir Al-Hashimi for his appreciable advice, important comments, support and encouragement.

Funding

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

Authors' contributions

All authors contributed toward data analysis, drafting and revising the paper and agreed to be responsible for all the aspects of this work.

Conflict of Interest

We have no conflicts of interest to disclose.

ORCID

Zahraa Dhahir: https://www.orcid.org/0000-0002-3724-3378

References

 British Pharmacopoeia. 7th edition. The Stationery office, Londone. 2012 [Publisher]
 Katzung B.G., Basic and Clinical Pharmacology. 2nd ed., Appleton and Lange.
 1987 [Crossref], [Google Scholar], [Publisher]
 British Pharmacopoeia. London. Her Majesty's Stationery Office. 2007 [Google Scholar] [4]. Turkey N.S., Al-Awadie, Ahmed F.K., *J. Kerbala uni.*, 2013, **9**:153 [Google Scholar], [Publisher]

[5]. Turkie N.S., Al-Awadi, Munshid H.Q., *J. Pharm.*

Sci., 2019, **11**:1563 [Google Scholar], [Publisher]

[6]. Nagham S.T., Mustafa K.K., *Int. J. Res. Pharm. Chem.*, 2016, **6**:759 [Publisher]

[7]. Al-Awadie N.S.T., Ibraheem M.H., , *Int. J. Res. Pharm.*, 2016, **6**:230 [Google Scholar]

[8]. Al-Awadie N.S.T., Khudhair A.F., *Iraqi J. Sci.*, 2014, **55**:12 [Google Scholar], [Publisher]

[9]. IRQ. PATENT, No. 3884, Shakir I.M.A., Al-Awadie N.S.T., ISNAG Fluorimeter For Fluorescence Measurement At The Visible Region Of Electromagnetic Radiation Using Mercury Tube Lamp As A Source For Irradiation For Continuous Flow Analysis Using Solar Cells. *Iraq, International classification*, 2014, H01J61/00

[10]. Al-Awadie N.S.T., lbraheem M.H., *Int. J. Res. Pharm. Chem.*, 2016, **6**:891 [Google Scholar]

[11]. Jordan W., van Barneveld H., Gerlich O., Kleine-Boymann M., Ullrich J., Phenol. *Ullmann's encyclopedia of industrial chemistry*. 2000. [Crossref], [Google Scholar], [Publisher]

[12]. Carey F.A., *Organic Chemistry*. 11rd Ed., McGraw-Hill, New York.1996 [Publisher]

[13]. Murdoch J., Barnes J.A., *Statistical Tables: for Science, Engineering, Management and Business Studies*. Macmillan International Higher Education. 1986 [Google Scholar], [Publisher]

[14]. Miller J.M., Miller J.C., *Stastical and chemometric for analytical chemistry*. Fifth Edition. Person eduction limited 2005 [Google Scholar]

[15]. British pharmacopoeia, Published by the stationary office London. 2003 [Crossref], [Publisher]

[16]. *The British Pharmacopoeia Commission Secretariat.* . Part of the Medicines and Healthcare products Regulatory Agency (MHRA). British Pharmacopoeia, Her Majesty's Stationery Office, London, UK2009 [Google Scholar]

HOW TO CITE THIS ARTICLE

Zahraa Dhahir. Nagham Shakir Turkie. New Turbidimetric Method for Determination of Cefotaxime Sodium in Pharmaceutical Drugs Using Continuous Flow Injection Manifold Design with CFTS- Vanadium Oxide Sulfate System, *Chem. Methodol.*, 2022, 6(2) 91-102 DOI: 10.22034/chemm.2022.2.2 URL: http://www.chemmethod.com/article 140435.html