Journal of Medicinal and Chemical Sciences 5 (2022) 1018-1025



Journal of Medicinal and Chemical Sciences

Journal homepage: <u>http://www.jmchemsci.com/</u>



### **Original Article**

## **RP-LC Method in Gradient mode for Combined Quantification of Metformin, Linagliptin, and Empagliflozin in Combination Tablets**

## C. Hazarathaiah Yadav 💿 , A. Mallibabu\* 💿

Department of Chemistry, Vel Tech Rangarajan Dr. Sagunthala R&D Institute of Science and Technology, Avadi, Chennai-600062, India

#### ARTICLE INFO

#### Article history

Received: 2022-02-17 Received in revised: 2022-03-11 Accepted: 2022-03-15 Manuscript ID: JMCS-2202-1423 Checked for Plagiarism: Yes Language Editor: Dr. Fatimah Ramezani Editor who approved publication: Prof. Dr. Hassan Karimi-Maleh


#### DOI:10.26655/JMCHEMSCI.2022.6.15

K E Y W O R D S Metformin Linagliptin Empagliflozin Gradient-RP-HPLC Fixed dosage forms

### ABSTRACT

It has been developed and validated an accurate, sensitive, precise, quick, and gradient reverse phase HPLC (RP-HPLC) technique for combined quantification of Metformin, Linagliptin, and Empagliflozin in raw materials and pharmaceutical combined tablets. With acetonitrile as the organic solvent, the best separation was achieved on a 250 mm x 4.6 mm i.d., 5µparticle size Inertsil®-Octadecyl-silyl-3V-Reverse-Phase-C18-column with acetonitrile as the non-polar-modifier and Dipotassium mono-hydrogen phosphate buffers system [0.03M] in water with pH 3.0 adjusted with orthophosphoric acid (0.1 %v/v) in the gradient mode of elution as an elution solvent at a speed of 1.0 mL.min<sup>-1</sup>. UV detection was at 230-nm. Metformin had a retention time of 2.421 minutes, Linagliptin's was 8.187 minutes, and Empagliflozin's was 11.71 minutes. With a correlation coefficient of about 0.9999, the peak-response was obtained as a function of concentration over the range of 80-480 mcg/mL for metformin, 0.4-2.4 mcg/mL for linagliptin, and 0.8-4.8 mcg/mL for empagliflozin. Metformin, Linagliptin, and Empagliflozin were shown to have a percentage assay of 99.92, 99.72, and 99.74, respectively. Metformin, Linagliptin, and Empagliflozin, each have a limit of detection of 0.04 g/mL, 0.004 g/mL, and 0.004 g/mL, respectively. Metformin, Linagliptin, and Empagliflozin, each have a limit of quantification (LOQ) of 0.12 ng/mL, 0.012 g/mL, and 0.012 g/mL, respectively. The presence of excipients in the formulation had no effect on the assay method. The procedure is appropriate for usage in QC- laboratories since it is quick and precise.

### GRAPHICAL ABSTRACT



\* Corresponding author: A. Mallibabu
⊠ E-mail: Email: <u>mallibabua@gmail.com</u>
© 2022 by SPC (Sami Publishing Company)

#### Introduction

Empagliflozin Na-glucose-coinhibits the transporter system that can be taken orally (SGLT2). The molecular weight is 450.91D and the chemical formula is C<sub>23</sub>H<sub>27</sub>ClO<sub>7</sub>. Empagliflozin is a non-hygroscopic white to yellowish powder [1-3]. It's soluble in water, C<sub>2</sub>H<sub>5</sub>OH, and acetonitrile; it's soluble in 1:1 proportion of CH<sub>3</sub>CN-H<sub>2</sub>O; and essentially not soluble in toluene<sup>4</sup>. Linagliptin is given by mouth, which inhibits the di-peptidylpeptidase-4 (DPP-4) [4]. It is soluble in water, methanol, ethanol, and acetonitrile, as well as soluble in 50 percent acetonitrile/water, and essentially insoluble in toluene<sup>4</sup> Linagliptin is an orally administered dipeptidyl peptidase-4 (DPP-4) inhibitor. The chemical formula is C<sub>25</sub>-H<sub>28</sub>-N<sub>8</sub>-O<sub>2</sub>, and the molecular weight is 472.54. Linagliptin is soluble in methanol, ethanol very sparingly, isopropanol only very faintly, and acetone only very weakly. Metformin-HCl is a crystalline whiteto-off-white powder with a molecular weight of 165.63D and a molecular formula of  $C_{4-}H_{11-}N_5\bullet HCl$ . Metformin-HCl is soluble in water but almost insoluble in acetone, pet-ether, and chloroform. The pKa-of metformin is 12.4 and metformin hydrochloride in a 1% aqueous solution has a pH of 6.68. Trijardy® is a novel FDC for diabetes therapy which combines extended-release metformin (metformin-XR) with empagliflozin, an SGLT2 inhibitor, and linagliptin, a DPP-4 inhibitor. Each TRIJARDY XR film coated tablet contains an extended-release metformin-HCl central tablet covered with the empagliflozin as fast-released pharmacological ingredient. In the literature, RP-HPLC approach has been published for measuring the potency of metformin and other oral antidiabetic drugs [5-8]. Nevertheless, no techniques for combined determination of metformin-HClwith-linagliptin & empagliflozin have been published in oral fixed dosage form. Furthermore, no official or preliminary monograph on this combination of analytes has been published in any of the compendial pharmacopoeias [9]. The goal of this study was to develop an accurate and efficient RP-HPLC method to estimate a new combination of antidiabetic drugs in fixed dosage forms for oral administration. The validation of the devised approach is also addressed in this study, as per ICH standards [12].

#### **Martials and Methods**

• Metformin, 99%, Linagliptin of 99% and Empagliflozin of 99% pure are acquired from Sigma-Aldrich Chemicals, Mumbai, India.

• Rankem-Fine-Chemicals of HPLC- Grade-Acetonitrile.

• Rankem-Fine-Chemicals AR\_grade: DiPotassium-monohydrogen-phosphate (dibasic, K<sub>2</sub>HPO<sub>4</sub>) [0.03M].

• *Ortho*-H<sub>3</sub>PO<sub>3</sub>, 85% (v/v) obtained from Quligen-Fine chemicals.

• Chromatographic-Grade water.

#### *Chromatographic-Instrument*

Quantitative HPLC was carried out on a Shimadzu Prominence high-performance liquid chromatograph with a PDA detector module, which included an automated injector with a 20-l injection volume and a quadra-pump. The utilized column was a Reverse-Phase-Inertsil-Octa-Decyl-S-3V-C<sub>18</sub> column (250 mm x 4.6 mm internal diameter with particle size 5 µm). Spin-chrome Software was installed on the HPLC equipment. The column temperature 40 °C was adjusted and eluted over 20.0 minutes at a mobile solvent speed of 1.0 mL.min<sup>-1</sup> under linear gradient conditions. The organic modifier is acetonitrile, while the mobile phase is water with Dipotassium hydrogen phosphate [0.03M] and a pH of 3.0 adjusted with ortho-phosphoric acid (0.1 percent v/v) (0.1% v/v). It was degassed and filtered via 0.45 m Nylon membrane filters before use. For three analytes, UV detection at 230 nm was used as isobestic wavelength detection with a PDA detector.

#### Preparation of Primary Standard Drug Solutions

To make the primary standard stock solution, 400mg of Metformin, 2 mg of Linagliptin, and 4 mg of Empagliflozin were dissolved in a volumetric flask (100 mL) with 20 mL of diluent (1:1 v/v CH<sub>3</sub>CN:H<sub>2</sub>O), sonicated for 15 minutes, and then brought up to 100mL with diluents to get the primary standard stock solution containing 4000  $\mu$ g.mL<sup>-1</sup> of Metformin, 20  $\mu$ g.mL<sup>-1</sup> of Linagliptin and 40  $\mu$ g.mL<sup>-1</sup> of Empagliflozin.

1019 | P a g e

Preparation of Working Standard Drug Solution In a 50 mL volumetric flask containing 400  $\mu$ g.mL<sup>-1</sup> Metformin, 2 $\mu$ g.mL<sup>-1</sup> Linagliptin, and 4  $\mu$ g.mL<sup>-1</sup> Empagliflozin, the working standard solution was produced up to 50 mL with diluent (1:1 v/v CH<sub>3</sub>CN:H<sub>2</sub>O) from 5 mL of the above-mentioned primary-stock solution.

### Sample Preparation

Twenty Trijardy-XR® combination tablets, which contain an extended-release metformin-HCl central tablet covered by the empagliflozin and linagliptin as fast-released pharmacological ingredient components are available in 10 mg empagliflozin/5 mg linagliptin/1000 mg metformin-HCl extended-release combined tablet dosage forms, were accurately weighed, and the mean weight was determined, as well. An equivalent-quantity of the homogeneous powder equivalent to 400 mg metformin, 4 mg empagliflozin, and 2 mg linagliptin was transferred, and 50 mL 0.02 M KH<sub>2</sub>PO<sub>4</sub> and 0.02 M  $K_2$ HPO<sub>4</sub> added. CH<sub>3</sub>OH:H<sub>2</sub>O at ratio of 90:10 (v/v) was added as diluent, sonicated for 30 minutes, diluted to 100 mL and pH adjusted to 7.0 with ortho-H<sub>3</sub>PO<sub>3</sub> in water. The resulting solution was miscible completely by sonication for 30 min, diluted up to 100 mL with ortho-phosphoric acid. Metformin 400 µg.mL<sup>-1</sup>, Linagliptin 2 µg mL<sup>-1</sup>, and Empagliflozin 4  $\mu$ g mL<sup>-1</sup> are all present in the sample solution. The individual linear regression equations were used to determine the amounts of metformin, empagliflozin, and linagliptin in teritiary mixture of sample solution form fixeddosage-forms.

### Linearity

Aliquots of Metformin, Linagliptin, and Empagliflozin working stock solutions were placed in various 10mL volumetric flasks and made the volume up to the 10mL with the mobile phase, yielding in final strengths of 80-480 μg.mL<sup>-1</sup>, 0.4-2.4 μg.mL<sup>-1</sup>, and 0.8-4.8 μg.mL<sup>-1</sup>, respectively (Table 1). The peak areas and retention times of each of these drug solutions (loaded at 10  $\mu$ L) were measured three times in the column. Using a PDA-detector set at 230 nm, a linearity-graph was generated by plotting peak areas versus Metformin, Linagliptin, and Empagliflozin concentrations in µg-mL<sup>-1</sup>.

#### Accuracy

The approach's accuracy was found by evaluating the drugs' recovery by using the standard-spiking method. To assess if the analytes contained in the positive or negative formulation caused interventions, known amounts of each drug equivalent to 10 percent standard drug solution were added to 80 percent, 100 percent, and 120 percent of the target test concentrations a formulation mixture. Each set-of-addition was replicated three times at each dilution level. The results are compared to a competent reference standard after extraction of sample preparation. The percentage of analytes re-covered by the assay was used to assess the accuracy. Table 4 shows the results of accuracy investigations on standard solution and process-related impurity; the recovery measurements suggest that the procedure was accurate.

### Precision

Quality-control samples in 100 % (w/v) dilution were used to assess intraday and interday precision. On the same day, six replicates of the target concentrations were examined for intraday variation, and six replicates were examined for inter-day variation on three different days. The method's repeatability is indicated by the low RSD value (1%) (Table 4).

### Limits of Detection and Quantification

The method's LOD was set at the lowest concentrations of active pharmaceutical components with a signal-to-noise (S/N) ratio of around 3 (LOD). The lowest active therapeutic medication concentrations which can be assessed with acceptable precision and accuracy while maintaining a signal-to-noise (S/N) ratio of roughly 10 (LOQ).

### Method Applicability

Our study group evaluated the newly created method by applying it to pharmaceutical tablets for estimating Metformin, Linagliptin, and Empagliflozin.

### Method Validation Tests

Method precision (RSD, percent), method accuracy (recovery percent & %RSD), linear range (r<sup>2</sup>), and LOD & LOQ were explored as recommended method validation characteristics.

1020 | P a g e

Babu A.M., et al. / J. Med. Chem. Sci. 2022, 5(6) 1018-1026

#### **Result and Discussion**

### Linearity

With a correlation coefficient of 1.0, the graph of chromatographic-peak areas of all analytes versus respective concentrations was indicated to be linear in the band of 80-480  $\mu$ g. mL<sup>-1</sup> for Metformin, 0.4-2.4  $\mu$ g.mL<sup>-1</sup> for Linagliptin, and 0.8-4.8  $\mu$ g.mL<sup>-1</sup> for Empagliflozin (Table 1). The least

square fit data of linear regression analysis was derived from the measurements given in Table 2. Metformin's regression formula is y = 1E+06x, Linagliptin's is y=122817x, and Empagliflozin's is y = 187710x. Table 2 presents the regression parameters for this technique that include slope, intercept, and % RSD. These findings suggest that there was a significant correlation.

<b>Table 1:</b> Evaluation of precision with-in-day and day-to-day analysis	

With-in-day Precision study of 100% dilution containing Metformin (400 µg/Ml), Linagliptin (2 µg/mL) and Empaglifozin (4 µg/mL)						Day-to-Day Precision study of 100% dilution containing – Metformin (400 μg/mL), Linagliptin (2 μg/mL) and Empaglifozin (4 μg/mL)							
	Metform	in	Lina	gliptin	Empag	gliflozin	Met	tformin	Lina	gliptin	Empa	gliflozin	
S. No	R.T	Peak-Area	R.T	Peak- Area	R.T	Peak- Area	R.T	Peak-Area	R.T	Peak- Area	R.T	926267	
1	2.421	5735858	8.187	622809	11.716	935541	2.425	5771972	8.194	607055	11.693	927709	
2	2.422	5738972	8.189	606552	11.713	922964	2.426	5773228	8.195	608105	11.696	928509	
3	2.422	5753564	8.189	605189	11.702	921717	2.426	5778001	8.196	608723	11.692	929952	
4	2.424	5743597	8.191	605106	11.700	922997	2.426	5779250	8.200	609552	11.699	930743	
5	2.425	5747941	8.194	605885	11.704	923517	2.426	5784303	8.198	610453	11.693	931079	
6	2.425	5751030	8.194	606724	11.698	923612	2.426	5792709	8.193	610354	11.685	929043.2	
Average	2.423	5745160.3	8.191	608710.8	11.705	925058	2.426	5779910.5	8.196	609040.3	11.693	1877.9	
Std. Dev	0.0017	6929.4	0.0028	6939.0	0.0073	5179.8	0.0042	7679.2	0.0024	1334.2	0.0046	0.2	
% RSD	0.060	0.1	0.033	1.1	0.061	0.6	0.019	0.1	0.031	0.2	0.040	926267	

Table 2: Regression analysis & Operating-System Suitability Results

Metformin	Linagliptin	Empagliflozin
2.421	8.187	11.716
1.454	1.399	1.090
5735858	622809	935541
78.65	8.53	12.82
3033.879	17953.186	31545.673
0.000	27.439	13.884
80-480	0.4-2.4	0.8-4.8
0.04	0.004	0.004
0.12	0.012	0.012
0.998	0.998	0.998
99.92	99.32	99.74
	2.421 1.454 5735858 78.65 3033.879 0.000 80-480 0.04 0.12 0.998	2.421     8.187       1.454     1.399       5735858     622809       78.65     8.53       3033.879     17953.186       0.000     27.439       80-480     0.4-2.4       0.04     0.004       0.12     0.012       0.998     0.998

### Accuracy

Individual recovery of analyte at 80 %-dilution level on w/v basis, 100 %-dilution level on w/v basis and 120 %-dilution level on w/v basis of prescribed concentrations was 98.1 percent to 102.45 percent, demonstrating the method's accuracy. The RSD was usually less than 1% in these data, demonstrating that the technique seems to be very accurate and generates consistent results (Table 3).

Amount-of Metformin (μg/mL)	Chromatogram	Amount-of Linagliptin (µg/mL)	Chromatogram	Amount-of Empagliflozin in µg/mL	Chromatogram
80	Peak- Area	0.4	Peak- Area	0.8	Peak- Area
160	1196011	0.80	124271	1.6	189169
240	2351071	1.20	244392	2.4	370631
320	3525071	1.60	373497	3.2	562549
400	4690599	2	492631	4.0	750593
480	5756301	2.4	609699	4.8	939968

Table 3: Summery of the standard calibration Curve for Linearity experiment

1021 | P a g e

Babu A.M., et al. / J. Med. Chem. Sci. 2022, 5(6) 1018-1026

#### Precision

Table 4 summarizes the intra-day and inter-day fluctuation in precision analysis. The method's repeatability is indicated by the low RSD value (lessthan-1%). These results indicate that the approach has a high level of precision and repeatability, both within a single analytical run and across multiple runs (Table 4).

	Standard-addition with 80%w/v of the Working Standard			Standard-addition with 100%w/v of the Working Standard			Standard-addition with 120% (w/v) of the Working Standard		
Accuracy parameter	Empagliflozin (3.2 μg/mL)	Linagliptin (1.60 µg/mL)	Metformin (320 µg/mL)	Empagliflozin (4.0 μg/mL)	Linagliptin (2.0 µg/mL)	Metformin (400 µg/mL)	Empagliflozin (4.8 μg/mL)	Linagliptin (2.4 µg/mL)	Metformin (480 µg/mL)
Amount added	0.40	0.20	40	0.4	0.2	80	0.4	0.2	80
Amount Found	3.57	1.78	358	4.36	2.19	479	5.22	2.41	558
% Recovery	99.16	98.88	99.44	99.09	99.54	99.79	100.2	99.58	99.64
% RSD.	0.02	0.04	0.15	0.28	0.54	0.22	0.62	0.74	0.82

Table 4: Accuracy	evaluation by Spike-analysis method
-------------------	-------------------------------------

### Robustness

At three different levels, -ve2, 0, and +ve2, the influence of minute but intentional alterations in the separation parameters was explored. The experimental settings were purposely changed at three distinct levels to determine the robustness of this approach, and the RT and chromatographic responses were evaluated. The effect was

observed by modifying one factor at a time. The RT and the method response were not affected by changing the stationary phase, pH of elution solvents by 0.2 units (pH=3.2 and pH=2.8), or mobile phase flow rate by 1.0 mL.min-1 (0.8 and 1.2 mL.min<sup>-1</sup>, demonstrating that the procedure was robust. Table 5 summarises the findings.

<b>Table 5:</b> The method's Robustness can be tested by changing the chromatographic settings (n=3, 100%w/v
Working dilution contains Metformin-400 $\mu$ g/mL, Linagliptin- 2 $\mu$ g/mL and Empagliflozin-4 $\mu$ g/mL)

	-									
	Study of Pe	ak areas with l	Different HPLC	Study of Pe	eak areas with	Mobile Phase	Study of Peak areas with Mobile Phase			
S.No.		Column			pH increase	2	pH decrease			
	Metformin	Linagliptin	Empagliflozin	Metformin	Linagliptin	Empagliflozin	Metformin	Linagliptin	Empagliflozin	
1	6035770	633834	971248	5658854	545876	837626	5289148	639916	979780	
2	6045001	634525	971486	5661336	542994	828083	5298146	641478	980233	
3	6047370	635116	972876	5660917	548447	826794	5302371	642333	981138	
Mean	6042713	634492	971870	5660369	545772	830834	5296555	641242	980384	
%	0.101	0.101	0.090	0.023	0.500	0.712	0.128	0.191	0.070	
RSD				0.010	0.000	0.011	0.120	0.171	0.070	
	Study of Re	etention times	with Different	Study of Retention times with Mobile			Study of Retention times with Mobile			
S.No.	No. HPLC Column			Phase pH increase			Phase pH decrease			
	Metformin	Linagliptin	Empagliflozin	Metformin	Linagliptin	Empagliflozin	Metformin	Linagliptin	Empagliflozin	
1	2.420	8.166	11.732	2.257	7.943	10.973	2.345	8.128	11.718	
2	2.417	8.162	11.722	2.259	7.937	10.972	2.347	8.126	11.713	
3	2.418	8.163	11.728	2.257	7.930	10.970	2.347	8.124	11.718	
Mean	2.418	8.164	11.728	2.257	7.937	10.972	2.346	8.126	11.716	
% RSD	0.044	0.029	0.044	0.041	0.080	0.016	0.061	0.022	0.027	

### Limit-of-Detection & Limit-of-Quantifications

The limit-of-detection (LOD) for Metformin is  $0.04\mu g/mL$ , and for Linagliptin and Empagliflozin was found to be  $0.004 \ \mu g.mL^{-1}$ . The limit-of-quantification (LOQ) for Metformin is  $0.12 \ \mu g.mL^{-1}$ 


<sup>1</sup>, and for Linagliptin and Empagliflozin was found to be 0.012  $\mu$ g.mL<sup>-1</sup>. These numbers illustrate the method's high sensitivity, which is essential in most investigations, as well as the fact that it can

1022 | P a g e

be used to detect and quantify analytes over a wide concentration range.

#### Specificity

The RTs for Metformin, Linagliptin, and empagliflozin were determined to be 2.42 minutes for Metformin, 8.18 minutes for Linagliptin, and 11.72 minutes for empagliflozin, according to the representative chromatogram given in Figure 1. When the pharmaceutical tablet matrices were evaluated, no indication of excipient interference signals was observed in the respective RTs of the chromatogram. It indicates that the analytes were not disturbing the probable merging peaks. As a result, this technique can be employed with certainty.



**Figure 1:** Chromatogram of Metformin-400µg/mL, Linagliptin-2µg/mL and Empagliflozin- 4µg/mL analyzed by the optimized Gradient RP-HPLC method




Figure 2: Linearity graphs of metformin for dilution of standard solutions




Figure 3: Linearity graphs of Linagliptin for dilution of standard solutions

<sup>1023 |</sup> P a g e

Babu A.M., et al. / J. Med. Chem. Sci. 2022, 5(6) 1018-1025




Figure 4: Linearity graphs of empagliflozin for dilution of standard solutions

A gradient RP- HPLC procedure for assaying the active ingredients was developed due to lack of an easy, reproducible, and quick-to-use method for the determination of Metformin, Linagliptin, and Empagliflozin concentrations in formulary matrices. We examined the effect of various HPLC technique variables on the final result of the study to optimize the chromatographic parameters, various proportions of CH<sub>3</sub>OH-H<sub>2</sub>O, CH<sub>3</sub>CN-H<sub>2</sub>O, and CH<sub>3</sub>CN-KH<sub>2</sub>PO<sub>4</sub> buffer were tested. After several early investigatory tests, CH<sub>3</sub>CN-KH<sub>2</sub>PO<sub>4</sub> buffer system [pH 3.0], it was chosen over other mobile phases because it resulted in improved resolution of pharmacologically active components. This procedure gives the good separation of analytes after multiple exploratory & investigatory trail runs. All three active pharmaceutical analytes had an excellent UV sensitivity and were interference-free at 230 nm. The analyte peaks were highly defined and without any incidence of tailing under these conditions. The aforementioned set of conditions were chosen for additional validation after considering the entire body of data acquired from this extensive study.

## Conclusion

In this study, an efficient and commonly available HPLC method was devised for analyzing Metformin, Linagliptin, and Empagliflozin in pharmaceutical matrices. This method's key advantages are its significantly reduced run times, ease of use, and of operation. All of these features are critical in operation, especially when analyzing a large number of samples. The validation experiments demonstrated that the procedural approach has a large calibration concentration range, adequate precision, accuracy, and practically reliable sensitivity. The method can be used for regular analysis in formulation QCstudies and allows for a straightforward, selective, sensitive, and specific assessment of Metformin, Linagliptin, and Empagliflozin.

### Acknowledgments

The authors are grateful to the Management of Vel Tech University and commercial organization of Vishnu Chemicals Ltd. Hyderabad for their encouragement and providing the necessary facilities to carry out the entry research work. **Funding** 

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

## Authors' contributions

All authors contributed toward data analysis, drafting and revising the paper and agreed to responsible for all the aspects of this work.

### **Conflict of Interest**

We have no conflicts of interest to disclose.

## **ORCID**:

C. Hazarathaiah Yadav <u>https://orcid.org/0000-0002-5124-879X</u> A. Mallibabu <u>https://orcid.org/0000-0001-8597-9606</u>

1024 | P a g e

Babu A.M., et al. / J. Med. Chem. Sci. 2022, 5(6) 1018-1025

### References

[1]. Yadav C.H., Babu A.M., Development and Validation of Head-space Gas Chromatographic Method in Tandem with Flame ionized detection for the determination of Residual solvents in Simeprevir API Synthesis. *Research Journal of Pharmacy and Technology*, 2021, **14**:5175 [Crossref], [Google Scholar], [Publisher]

[2]. Mallibabu A., Yadav C.H., Pathi P.J., Development and validation of a stability indicating method by ultra performance liquid chromatography for the determination of vancomycin drug in vancomycin injection, *World Journal of Pharmaceutical Research*, 2022 [Publisher]

[3]. Hazarathaiah, C., The Estimation of Mepivacain Hydrochloride Injection Liquid. International Journal of Engineering and Techniques, 2018, **4**:628 [PDF], [Google Scholar]

[4]. Bailey C.J., Day C., Fixed-dose single tablet antidiabetic combinations. Diabetes, *Obesity and Metabolism*, 2009 **11**:527 [Crossref], [Google Scholar], [Publisher]

[5]. Ayoub B.M., UPLC simultaneous determination of empagliflozin, linagliptin and metformin. *RSC advances*, 2015 **5**:95703 [Crossref], [Google Scholar], [Publisher]

[6]. Shafi M.S.S., Begum A., Saradhi N., Bioanalytical method development and validation of linagliptin in plasma through LC- MS/MS. *International Journal of Bioassays*, 2014, **3**:3146 [Google Scholar]

[7]. Shah P.A., Shrivastav P.S., George A., Mixedmode solid phase extraction combined with LC-MS/MS for determination of empagliflozin and linagliptin in human plasma. *Microchemical Journal*, 2019, **145**:523 [Crossref], [Google Scholar], [Publisher]

[8]. Donepudi S., Achanta S., Validated HPLC-UV method for simultaneous estimation of linagliptin and empagliflozin in human plasma. *International Journal of Applied Pharmaceutics*, 2018, **10**:56 [Crossref], [Google Scholar], [Publisher]

[9]. Abdel-Ghany M.F., Abdel-Aziz O., Ayad M.F., Tadros M.M., New LC–UV methods for pharmaceutical analysis of novel anti-diabetic combinations. *Acta Chromatographica*, 2017, **29**:448 [<u>Crossref</u>], [<u>Google Scholar</u>], [<u>Publisher</u>]

[10]. Mabrouk M.M., Soliman S.M., El-Agizy H.M., Mansour F.R., A UPLC-DAD method for simultaneous determination of empagliflozin and three related substances in Standard-addtion human plasma. *BMC Chem.*, 2019, **13**:83 [<u>Crossref</u>], [<u>Google Scholar</u>], [<u>Publisher</u>]

[11]. Bhole R.P., Wankhede S.B., Pandey M., Stability indicating HPTLC method for simultaneous estimation of empagliflozin and linagliptin in pharmaceutical formulation. *Analytical Chemistry Letters*, 2017, **7**:76 [Crossref], [Google Scholar], [Publisher]

### HOW TO CITE THIS ARTICLE

C. Hazarathaiah Yadav, A. Mallibabu. RP-LC Method in Gradient mode for Combined Quantification of Metformin, Linagliptin, and Empagliflozin in Combination Tablets, *J. Med. Chem. Sci.*, 2022, 5(6) 1018-1025 <a href="https://doi.org/10.26655/JMCHEMSCI.2022.6.15">https://doi.org/10.26655/JMCHEMSCI.2022.6.15</a> URL: <a href="https://www.jmchemsci.com/article149694.html">https://doi.org/10.26655/JMCHEMSCI.2022.6.15</a> URL: <a href="https://www.jmchemsci.com/article149694.html">https://www.jmchemsci.com/article149694.html</a>

1025 | P a g e