

EDITORIAL

The Most Appropriate Reconstruction Method Following Giant Cell Tumor Curettage: A Biomechanical Approach

Azadeh Ghouchani, MSc; Mohammad H. Ebrahimzadeh, MD; Gholamreza Rouhi, PhD

Research performed at the Department of Biomedical Engineering, Amirkabir University of Technology, Tehran, Iran

Giant cell tumor (GCT) is a primary and benign tumor of bone, albeit locally aggressive in some cases, such as in the epi-metaphyseal region of long bones, predominantly the distal end of femur and proximal end of tibia (1). There are a variety of treatments for a bone affected by GCT, ranging from chemotherapy, radiotherapy, embolization, and cryosurgery, to surgery with the use of chemical or thermal adjuvant (2). Even with advances in new chemotropic drugs, surgery is still the most effective treatment for this kind of tumor (3). The surgery often involves defect reconstruction following tumor removal (4). The aims of treatment are removing the tumor and reconstructing the bone defect in order to decrease the risk of recurrence, and restore limb function, respectively. To achieve these goals, reconstruction is usually accompanied with PMMA bone cement infilling (4). The high heat generated during PMMA polymerization in the body can kill the remaining cancer cells, and hence the chance of recurrence decreases (5). In addition, filling the cavity with bone cement provides immediate stability, enabling patients to return to their daily activities soon (6).

The major drawbacks of the technique of curettage and cementation is the high fracture risk, due to the early loading of the bone, and the insufficient fixation of the cement in the cavity (7). Hence, several

methods have been developed to fix the bone cement in order to prevent the postoperative fracture. Pattijn et al. packed the cement with a titanium membrane which was attached to the periosteum with small screws (7). The membrane can make early normal functioning of patients possible, since it partially restore the strength and stiffness of the bone. Cement augmentation with internal fixation is another method to decrease the risk of postoperative fractures (6, 8, 9). Common devices used as internal fixators include plates, screws, and pins, which provide immediate stability, as well as structural support for large defects (10). Nevertheless, there is still a controversy over the use of fixation devices, thus repairing large defects remains a challenging issue in orthopedic surgery, since whether or not these devices are really needed, and also what is the most proper device, is a place of debate (11).

Several in-vitro biomechanical studies have been done to evaluate the effect of internal fixation devices and different outcomes were resulted so far (8, 12-14). The use of Steinmann pin for cement augmentation following GCT curettage and cement infilling have shown promising results in proximal cadaveric tibia defects (9, 15). However, some other studies have not shown any biomechanical differences in terms of stiffness, energy to failure, yield point, and failure load between the specimens reconstructed with cement

Corresponding Author: Gholamreza Rouhi, Faculty of Biomedical Engineering, Amirkabir University of Technology, Tehran, Iran
Email: grouhi@aut.ac.ir; grouhi@uottawa.ca



THE ONLINE VERSION OF THIS ARTICLE
ABJS.MUMS.AC.IR

alone, and those which were cement augmented with pins (8, 16, 17). In an experiment in distal femur, cement augmented with each of two fixation devices including intramedullary Steinmann pin and crossed oriented screws were compared with the one with bone cement alone (6). The results showed a better performance, as expected, for the crossed screws cement augmented specimens, compared to specimen reconstructed with cement alone, under compressive cyclic load (6). Some in-vitro studies investigated the effectiveness of cement reinforcement with locking plates. The strength of bone, in the most persistent site of GCT, i.e. distal femur and proximal tibia, reconstructed with cement and fixed with locking plates, was compared with augmentation with crossed screws, or with Steinmann pins, which the results showed greater strength and stiffness of specimens reconstructed with cement reinforced with locking plates, compared to specimens reconstructed with cement and augmented with either crossed screws or Steinmann pins (14, 18).

Finite element (FE) method is a computer technique in which through using mesh generation a complex structure can be divided into small simple *elements* with *finite* numbers. The first application of FE technique in orthopedic biomechanics goes back to early 1970's (19). FE modeling and analysis have now wide applications in orthopedic surgery to tackle and solve numerous problems ranging from bone remodeling process prediction under different loads, to selecting the best fixation device for a specific bone trauma or fracture (20-25). In regard to GCT surgery, there are a few studies which employed FE approach combined with experimental tests to find out the fracture load, as well as the best fixation device for cement augmentation (26-29). A patient specific FE study was performed to select the most appropriate reconstruction method among: cavity with no filler; filling the cavity with bone cement alone; and bone cement augmented with locking plates following GCT curettage in distal femur, which showed a greater stiffness for the model having locking plates, compared to the two other models (28). Another group recently simulated GCT surgery through employing three dimensional heterogeneous FE models of bone, validated with in-vitro mechanical tests on cadavers, in order to calculate the fracture load, and its relation to the size and location of the tumors (29). They also compared their results with structural rigidity analysis and showed that both methods have great capability in predicting post-operative fracture load of a bone affected by GCT (29).

Due to the lack of consensus for the best reconstruction method, as well as disagreement among various researchers on the factors affecting long

bone fracture affected by GCT, and the multifactorial nature of the problem, which consists medical and engineering aspects, one can think of biomechanical methods for addressing this truly important concern. Recently, a biomechanical approach for finding a more comprehensive criterion for the problem of defect infilling, as well as for choosing the best fixation device for cement augmentation was introduced (30). The suggested approach comprised patients' specific factors including their weight, daily activities, and bone quality, as well as tumor-related factors including its size, location, and shape (30). Based on this approach, subject-specific quantitative CT (QCT)-based FE models of cadaveric distal femur bone reconstructed with cement were created and validated with in-vitro mechanical experiments (26, 27, 29). The model was capable of estimating bone strength close to that of in-vitro cadaveric data (26). The validated FE method, which is of course a non-invasive approach, can be applied prior to the surgery in order to help the surgeons decide about the most proper reconstruction method for each patient. In this way, GCT surgery can be mimicked and simulated virtually, and the outcomes can be analyzed either by a biomedical engineer, or a technician before the surgery is made by a surgeon. A QCT-scan of the affected limb by the tumor should be taken, then using suitable modeling software, such as Mimics or Simpleware, a 3-D model of the bone and tumor can be created. Based on the Hounsfield Units (HUs) of the images, material properties of bone can be assigned to the model, and cement material properties must be entered to the tumor region, since the tumor region in the surgery is filled with bone cement. The highest load the patient may experience during his/her daily activities should be applied on the model in order to see if it can withstand that level of stimulus. A schematic depiction of the FE modeling and analysis procedure can be seen in Figure 1.

Using the proposed model, depending on the gross geometry, and mass distribution of the long bone, as well as the size and location of the cavity, results of FE model can predict if cement infilling is needed, or besides cement infilling, internal fixators should be employed as well (30). It is hoped that, through using the biomechanical approach introduced here, see Figure 1, patients at high risk of post-operative fractures can be identified prior to the surgery, and prophylactic actions can be made to decrease the risk of post-operative fractures.

Acknowledgements

The authors would like to thank Amirkabir University of Technology, and Mashhad University of Medical Sciences for their support.




Figure 1. Various steps taken in making a voxel based finite element (FE) model of a long bone affected by GCT. The specimen is first scanned by QCT, in which a calibration phantom is used during the scanning procedure. DICOM images are imported into 3D modeling software for segmentation of the desired bones or parts. The 3D model is then meshed based on the QCT voxels. Using the calibration equation derived from known phantom tubes densities, as well as relations available in the literature for converting densities into mechanical properties, material properties can be assigned to each voxel of the model. Desired boundary and loading conditions are applied to the model, and the results in the form of stress, strain, and failure load can analyzed.

References

1. Cowan RW, Singh G. Giant cell tumor of bone: a basic science perspective. *Bone*. 2013; 52(1):238-46.
2. Singh S, Singh M, Mak I, Ghert M. Expressional analysis of GFP-Tagged cells in an in vivo mouse model of giant cell tumor of bone. *Open Orthop J*. 2013; 7:109-13.
3. Wu PF, Tang JY, Li KH. RANK pathway in giant cell tumor of bone: pathogenesis and therapeutic aspects. *Tumor Biol*. 2015; 36(2):495-501.
4. Fraquet N, Faizon G, Rosset P, Phillippeau J, Waast D, Gouin F. Long bones giant cells tumors: treatment by curettage and cavity filling cementation. *Orthop Traumatol Surg Res*. 2009; 95(6):402-6.
5. Wada T, Kaya M, Nagoya S, Kawaguchi S, Isu K, Yamashita T, et al. Complications associated with bone cementing for the treatment of giant cell tumors of bone. *J Orthop Sci*. 2002; 7(2):194-8.
6. Toy PC, France J, Randall RL, Neel MD, Shorr RI, Heck RK. Reconstruction of noncontained distal femoral defects with polymethylmethacrylate and crossed-screw augmentation: a biomechanical study. *J Bone Joint Surg Am*. 2006; 88(1):171-8.
7. Pattijn V, Gelaude F, Sloten JV, Van Audekercke R. Medical image-based preformed titanium membranes for bone reconstruction. *Med Imaging Syst Technol Methods General Anatomy*. 2005; 5:43-78.
8. Murray PJ, Damron TA, Green JK, Morgan HD, Werner FW. Contained femoral defects: biomechanical analysis of pin augmentation in cement. *Clin Orthop Relat Res*. 2004; 420(1):251-6.
9. Randall RL, Weenig KN, West JR, Johnston JO, Bachus KN. Durability and strength of Steinmann pin augmentation in cemented tibial defects. *Clin Orthop Relat Res*. 2002; 397(1):306-14.
10. Bickels J, Meller I, Malawer M. The biology and role of cryosurgery in the treatment of bone tumors. In: Malawer MM, Sugarbaker PH, editors. *Musculoskeletal cancer surgery*. Netherlands: Springer; 2001. P. 135-45.
11. Malawer MM, Bickels J, Meller I, Buch RG, Henshaw RM, Kollender Y. Cryosurgery in the treatment of giant cell tumor: a long term followup study. *Clin Orthop Relat Res*. 1999; 359(1):176-88.
12. Pan KL, Chan WH. Curettage and cementation in giant cell tumour of the distal tibia using polypropylene mesh for containment: A. *Malays Orthop J*. 2010; 4(2):51-3.
13. Toy PC, Arthur S, Brown D, Heck RK. Reconstruction of noncontained proximal tibial defects with divergent screws and cement. *Clin Orthop Relat Res*. 2007; 459(1):167-73.
14. Ugliarolo AD, Maceroli M, Beebe KS, Benevenia J, Patterson FR. Distal femur defects reconstructed with polymethylmethacrylate and internal fixation devices: a biomechanical study. *Orthopedics*. 2009; 32(8):561-67.
15. Bini S, Gill K, Johnston JO. Giant cell tumor of bone. Curettage and cement reconstruction. *Clin Orthop Relat Res*. 1995; 321(1):245-50.
16. Asavamongkolkul A, Pongkunakorn A, Harnroongroj T. Stability of subchondral bone defect reconstruction at distal femur: comparison between polymethylmethacrylate alone and steinmann pin reinforcement of polymethylmethacrylate. *J Med Assoc Thai*. 2003; 86(7):626-33.
17. Weiner M, Damron TA, Patterson FR, Werner FW, Mann KA. Biomechanical study of pins in cementing of contained proximal tibia defect. *Clin Orthop Relat Res*. 2004; 419(1):232-7.
18. Ruskin J, Caravaggi P, Beebe KS, Corgan S, Chen L, Yoon RS, et al. Steinmann pin augmentation versus locking plate constructs. *J Orthop Traumatol*. 2016; 17(3):249-54.
19. Brekelmans WA, Poort HW, Slooff TJ. A new method to analyse the mechanical behaviour of skeletal parts. *Acta Orthop Scand*. 1972; 43(5):301-17.
20. Kamal Z, Rouhi G. A parametric investigation of the effects of cervical disc prostheses with upward and downward nuclei on spine biomechanics. *J Mech Med Biol*. 2016; 16(2):1650092.
21. Chitsazan A, Herzog W, Rouhi G, Abbasi M. Alteration of strain distribution in distal tibia after triple arthrodesis: experimental and finite element investigations. *J Med Biol Eng*. 2017; 6(17):1-13.
22. Nourisa J, Rouhi G. Biomechanical evaluation of intramedullary nail and bone plate for the fixation of distal metaphyseal fractures. *J Mech Behav Biomed Mater*. 2016; 56(1):34-44.
23. Samsami S, Saberi S, Sadighi S, Rouhi G. Comparison of three fixation methods for femoral neck fracture in young adults: experimental and numerical investigations. *J Med Biol Eng*. 2015; 35(5):566-79.
24. Rouhi G, Vahdati A, Li X, Sudak L. A three dimensional computer model to simulate spongy bone remodelling under overload using a semi-mechanistic bone remodelling theory. *J Mech Med Biol*. 2015; 15(4):1550061.
25. Rouhi G, Tahani M, Haghighi B, Herzog W. Prediction of stress shielding around orthopedic screws: time-dependent bone remodeling analysis using finite element approach. *J Med Biol Eng*. 2015; 35(4):545-54.
26. Ghouchani A, Ebrahimzadeh MH, Rouhi G. Voxel-based finite element model of a reconstructed bone: simulating a bone tumor surgery. *J Bioeng Biomed Sci*. 2016; 6(5):47.
27. Ghouchani A, Ebrahimzadeh MH, Rouhi G. Predicting the risk of post-operative fracture of giant cell

- tumor using fully voxelised finite element models. 24th annual meeting of the Iranian orthopaedic association, September 26-30, Tehran, Iran; 2016.
28. Li J, Wodajo F, Theiss M, Kew M, Jarmas A. Computer simulation techniques in giant cell tumor curettage and defect reconstruction. *Comput Sci Engin.* 2013; 15(2):21-6.
29. Mosleh H, Rouhi G, Ghouchani A, Nourisa J, Bagheri N. Prediction of the fracture risk of reconstructed bone with cement using QCT based structural rigidity and finite element analysis. California: Orthopaedic Research Society Annual Meeting San Diego; 2017.
30. Ghouchani A, Rouhi G. The great need of a biomechanical-based approach for surgical methods of giant cell tumor: a critical review. *J Med Biol Engin.* 2017; 37(4):454-67.