¹DAVOOD BEIKI, ²ALI KHALAJ, ²REZA DOWLATABADI, ¹MOHAMMAD EFTEKHARI, ²MOHAMMAD HOSSEIN AL-SEYED HOSSEIN, ¹ARMAGHAN FARD, ¹BABAK FALLAHI, ³MOHAMMAD REZA KHOSHAYAND

¹Research Institute for Nuclear Medicine, Faculty of Medicine, ²Department of Medicinal Chemistry, ³Department of Food and Drug Analyses, Faculty of Pharmacy, Tehran University of Medical Sciences, Tehran, Iran

ABSTRACT

The urea breath test (UBT) which is carried out with ¹³C or ¹⁴C labeled urea is one of the most important non invasive methods for detection of *Helicobacter pylori* infection. Application of ¹³C-UBT is becoming increasingly popular because of its non radioactive nature which makes it suitable for diagnostic purposes in children and women of child bearing ages. While isotope ratio mass spectrometer (IRMS) is generally used to detect ¹³C in expired breath, this instrument is expensive and recently non dispersive isotope selective infrared (NDIR) spectroscopy which is a lower cost technique has been employed as a reliable counterpart for IRMS in small clinics. The aim of this study was to assess the validity of NDIR spectroscopy technique in Iranian population in comparison with histological examination, rapid urease test and ¹⁴C-urea breath test as gold standard. Seventy six patients with dyspepsia were underwent ¹³C-UBT and gold standard methods. The ¹³C-UBT showed 100% sensitivity, 97.3% specificity, 97.56% positive predictive value, 100% negative predictive value and 98.65% accuracy. On the basis of these results it could be concluded that ¹³C-UBT performed with NDIR spectroscopy is a reliable, accurate and non invasive diagnostic tool for detection of *Helicobacter pylori* infection in the Iranian population.

Keywords: *H. pylori*, ¹³C-urea breath test, Non dispersive isotope selective infrared spectroscopy.

INTRODUCTION

Helicobacter pylori has been recognized as a causative factor in various diseases including gastritis, duodenal ulcer, gastric ulcer, stomach cancer and Malt lymphoma (1-4). Several invasive and non invasive methods have been developed for detection of Helicobacter pylori infection. Invasive diagnostic methods require mucosal biopsy during endoscopy, followed by examination of the specimens by culture, rapid urease test or histological tests. Non invasive methods include antibody detection (serology), stool antigen test and urea breath test (5-6). While serology (ELISA) is simple and easy method to perform, it can neither differentiate active from remote infection, nor can it quickly confirm Helicobacter pylori eradication after treatment, since it remains positive for a long period despite successful eradication (7-8). Fecal antigen test is a practicable, non invasive alter-native with high sensitivity and specificity (6).

The urea breath test (UBT) is generally regarded as one of the best non invasive methods for diagnosis of *Helicobacter pylori* infection (9).

UBT is carried out using ¹³C or ¹⁴C labeled urea. In the presence of Helicobacter pylori, the labeled urea is metabolized by urease to yield labeled carbon dioxide which can be detected in the patient's expired breath as a marker of infection. While both isotopes seem to offer similar diagnostic accuracy, ¹⁴C-UBT is generally accepted as simpler, faster and cheaper technique. The main issue surrounding the use of ¹⁴C-UBT is the safety of ¹⁴C-urea. Being a radioactive isotope, concerns exist about its safe handling, as administration and disposal well as appropriateness of its usage in children and pregnant women. In contrast ¹³C is a non radioactive isotope that can be used safely for detection of Helicobacter pylori infection in children and women of child bearing age (10). The most commonly used method for the analysis of ¹³CO₂ in expired breath has been application of an isotope ratio mass spectrometer (IRMS) which is an extremely sensitive analytical device that measures ${}^{13}\text{CO}_2$ as the ${}^{13}\text{CO}_2/{}^{12}\text{CO}_2$ ratio (11). However this sophisticated device has the great disadvantage of high cost. The widespread

Correspondence: Davood Beiki, Research Institute for Nuclear Medicine, Tehran University of Medical Sciences, Shariati Hospital, North Kargar Ave., 14114, Tehran, Iran, E-mail: beikidav@sina.tums.ac.ir

demand for ¹³C-urea breath tests has led to development of cheaper and more practicable analytical systems such as non dispersive isotope selective infrared (NDIR) spectroscopy (10-12, 14-15) and laser assisted ratio analyzer (LARA) (10-13). On the basis of slightly different absorption spectra between ${}^{13}CO_2$ and ${}^{12}CO_2$ carbon dioxide molecules the isotope ratio can be determined using non dispersive isotope selective infrared (NDIR) spectrometer. In comparison with IRMS, NDIR spectrometer is a low cost device that preserves the precision of the test, hence there have been a growing number of reports of the ¹³C-UBT using a NDIR spectroscopy (11). The aim of this study was to assess the validity of NDIR spectroscopy technique in Iranian population in comparison with histological examination, rapid urease test and ¹⁴C-urea breath test as gold standard.

MATERIALS AND METHODS

¹³C-urea was purchased from Campro Scientific, Netherlands. The ¹³C-UBT was performed using a non dispersive infrared spectrometer (Heli FAN plus, Fischer ANalysen Instrumente GmbH, Germany). Breath samples were collected in aluminum plastic bags (Tecobag) purchased from Tesseraux GmbH, Germany.

Seventy six patients (40 female and 36 male; age of 18-66 years; median age of 39) who were referred for endoscopy because of digestive disorders to Shariati Hospital, Tehran University of Medical Sciences, Tehran, Iran were referred by specialists included in this study. The study was performed in the Research Institute for Nuclear Medicine, Tehran University of Medical Sciences, Tehran, Iran. Females who were pregnant, patients who had a history of gastric surgery and had taken bismuth and antibiotics within the last 4 weeks or proton pump inhibitors within the last 14 days were excluded from the study.

After obtaining written informed consent, all patients underwent upper gastrointestinal endoscopy with antral biopsies as well as ¹⁴C-urea breath test within one week (16).

Rapid urease test, histological examination and ¹⁴C-UBT were used as gold standard in this study. The microorganism was detected by Haematoxvlin and Eosin and modified Giemsa staining. Specimens were examined by expert pathologists who were blinded to all clinical information. Patients were considered to be Helicobacter pylori positive (HP+) and negative (HP-) when the results of two of the three tests were in agreement respectively. The ¹³C-urea breath test was carried out by a previously reported method (12).

After overnight fasting, patients were asked to drink a solution of 75 mg of ¹³C-urea in 200 ml orange juice. Breath samples were collected before and 30 minutes after drinking of the test solution in aluminum plastic bags with a volume capacity of 300 ml. The bags were immediately connected to the NDIR spectrometer and were analyzed within a few minutes. The results were presented as delta-over-baseline values (DOB) which indicate the change in the ${}^{13}CO_2$ to ${}^{12}CO_2$ ratio in the expired breath samples. The test was considered positive for Helicobacter pylori infection when the $\Delta \delta^{13}$ CO₂ value was greater or equal to 3.5 parts per thousand. The sensitivity, specificity, positive predictive value (PPV), negative predictive value (NPV), accuracy and associated confidence intervals (CI) for ¹³C-UBT were determined in comparison to the gold standard (17). Statistical analyses were performed with SPSS for windows statistical package (SPSS, Chicago, IL).

RESULTS AND DISCUSSION

A summary of demographic characteristics and endoscopic findings of the patients is presented in Table 1. Based on the gold standard examinations the number of true positive cases for *Helicobacter pylori* infection was 40 in this study and the remainder was considered as true negative cases.

Table 1. Clinical and endoscopic characteristics of the patients from the *Helicobacter pylori* (HP+) and (HP-) groups.

	HP (+)	(HP-)		
Female	21	19		
Male	19	17		
Endoscopic Findings				
Normal	10	13		
Peptic Ulcer	3	1		
Gastritis	8	11		
Esophagitis	4	7		
Gastritis + Duodenal Ulcer	4	0		
Gastritis + Esophagitis	8	3		
Gastritis + Duodenal Ulcer + Peptic Ulcer	3	0		
Duodenal Ulcer +Peptic Ulcer	0	1		

When compared to the gold standard, the ¹³C-UBT showed only one false positive result, leading to sensitivity, specificity, positive predictive value (PPV), negative predictive value (NPV) and accuracy of 100%, 97.3%, 97.56%, 100% and 98.65% (Confidence intervals 95%) respectively (Table 2). In the case of false positive result, the result of ¹⁴C-urea breath test was also positive which suggests due to probable sampling bias the invasive methods could have missed the infected patient (18).

Sensitivity	Specificity	PPV	NPV	Accuracy
(CI 95%)	(CI 95%)	(CI 95%)	(CI 95%)	(CI 95%)
100%	97.3%	97.56%	100%	98.65%

Table 2. Diagnostic performance of the ¹³C-urea breath test for the detection of *Helicobacter pylori* infection.

PPV= Positive Predictive Value, NPV= Negative Predictive Value, CI= Confidence Interval The sensitivity, specificity, positive predictive value, negative predictive value and accuracy were calculated according to literature (17).

Different studies have rated the sensitivity and specificity of this test at 90-97.8% and 96-100% respectively (12, 20). Although some studies have demonstrated that NDIR spectrometry is at least as accurate as mass spectrometry (15), some others have rated this technique inferior to IRMS (21), and as a result it is essential to validate this technique prior to extensive clinical application in ethnic population. In the present study, the results of the infrared spectroscopy method were compared with those obtained by a combination of biopsy methods and ¹⁴C-UBT in an Iranian population. Our results showed that ¹³C-UBT with NDIR spectrometry is 100% sensitive and 97.3% specific in comparison with the urease test, histological examination and ¹⁴C-UBT grouped together.

CONCLUSION

¹³C-UBT by NDIR spectrometry was found as a highly sensitive, specific and non invasive method for detection of Helicobacter pylori infection in an Iranian population. The low cost and technical simplicity are advantages of this technique, and make the method particularly effective for routine clinical practice.

ACKNOWLEDGMENT

We would like to thank the research council of Tehran University of Medical Sciences for financial support of this investigation.

REFERENCES

- Sipponen P, Hyvarinen H. Role of Helicobacter pylori in the pathogenesis of gastritis, peptic ulcer 1. and gastric cancer. Scand J Gastroenterol 1993; 28 (Suppl) 196: 3-6.
- Parsonnet J, Hansen S, Rodriguez L, Gelb AB, Warnke RA, Jellum E, Orentreich N, Vogelman JH, 2. Friedman GD. Helicobacter pylori infection and gastric lymphoma. N Engl J Med. 1994; 330(18):1267-71.
- Marshall BJ, Helicobacter pylori. Am J Gasteroenterol 1994; 89(Suppl 1): 116-128.
 Kuipers EJ, Thijis JC, Fasten HP. The prevalence of Helicobacter pylori in peptic ulcer disease. Aliment Pharmacol Ther 1995; 9: 59-69.
- 5. Atherton JC. Non-endoscopic tests in the diagnosis of Helicobacter pylori infection. Aliment Pharmacol Ther 1997; 11(Suppl 1): 11-20.
- 6. Gisbert JP, Pajares JM Stool antigen test for the diagnosis of Helicobacter pylori infection: a systematic review. Helicobacter 2004; 9(4):347-68.
- 7. Thijs Jc, Van Zwet AA, Thijs WJ, Oey HB, Karrenbeld A, Stellaard F, Luijt DS, Meyer BC, Kleibeuker JH. Diagnostic tests for *Helicobacter pylori*: a prospective evaluation of their accuracy, without selecting a single test as gold standard. Am J Gastroenterol 1996; 91: 2125-2129.
- 8. Walsh Jh, Peterson WL. The treatment of Helicobacter pylori infection in the management of peptic ulcer disease. N Engl J Med 1995; 333: 984-991.
- 9. Hunt RH, Thomson ABR. Canadian Helicobacter pylori consensus conference. Can J Gastroenterol 1998: 12: 31-41.
- 10. Savarino V, Vigneri S, Celle G. The urea breath test in the diagnosis of Helicobacter pylori infection. Gut 1999; 45(Suppl 1): 118-122.
- 11. Kato M, Saito M, Fukuda S, Kato C, Ohara S, Hamada S, Nagashima R, Obara K, Suzuki M, Honda H, Asaka M, Toyota T.¹³C-urea breath test, using a new compact nondispersive isotope- selective infrared spectrophotometer; Comparison with mass spectrometry. J Gastroenterol 2004; 39: 629-634.
- 12. Braden B, Caspary W.F, Lembcke B. Nondispersive infrared spectrometry for ¹³CO₂/¹²CO₂ measurements: A clinically feasible analyzer for stable isotope breath tests in gastroenterology. J Gastroenterol 1999; 37: 477-481.
- 13. Murnick DE, Peer BJ. Laser based analysis of carbon isotope ratios. Science 1994; 263: 945-947.

- 14. Haisch M, Hering P, Fuss W, Fabinski W. A sensitive isotope selective non dispersive infrared spectrometer for ¹³CO₂ and ¹²CO₂ concentration measurements in breath samples. Isopenpraxis Environ Health Stud 1994; 30: 247-251.
- 15. Koletzko S, Haisch M, Seeboth I, Braden B, Hengels K, Koletzko B, Hering P. Isotope selective non dispersive infrared spectrometry for detection of Helicobacter pylori infection with ¹³C-urea breath test. Lancet 1995; 345: 961-962.
- 16. Dowlatabadi Bazaz R, Khalaj A, Beiki D, Eftekhari M, Al-Seved Hosein MH, Khoshayand MR Microdose ¹⁴C urea breath test for the diagnosis of Helicobacter pylori: A survey in Iranian population. Daru 2005; 13(1): 6-10
- 17. Rosner B. Fundamentals of biostatistics. Boston: PWS-Kent Publishing Company; 1990.
- 18. Gonzalez P, Galleguillos C, Massardo T, Rivera M, Morales A, Smok G, Moyano L, Pimentel C, Alay R, Otarola S. Could the $[{}^{14}C]$ urea breath test be proposed as a gold standard for detection of Helicobacter pylori infection? Med Sci Monit 2003; 9: CR363-368.
- 19. Wong WM, Wong BCY, Wong KW, Fung FMY, Lai KC, Hu WHC, Yuen ST, Leung SY, Lau GKK, Lai CL, Chan CK, Go R, Lam SK 13C-urea breath test without a test meal is highly accurate for the detection of Helicobacter pylori infection in Chinese. Aliment Pharmacol Ther 2000; 14:1353-1358.
- 20. Chua TS, Fock KM, Teo EK, Ng TM, Validation of ¹³C-urea breath test for diagnosis of Helicobacter pylori infection in the Singapore population. Singapore Med J 2002; 43(8): 408-411.
- 21. Braden B, Schafer F, Caspary WF, Lembcke B. Nondispersive isotope-selective infrared spectroscopy: a new analytical method for ¹³C-urea breath tests. Scand J Gastroenterol 1996; 31(5):442-445.