Iranian Journal of Analytical Chemistry

Volume 9, Issue 1, March 2022 (1-14)

Original Research Article

مجله ایرانی شیمی تجزیه دوره ۹، شماره ۱، فروردین ۱۴۰۱ (۱۴–۱)

Application of Non-bilinear Voltammetric Data for the First-Order Excellence in Simultaneous Determination of Ascorbic Acid, Uric Acid, Acetaminophen and Noradrenalin in Interfering Media

Ghobad Mansori^{1*}, Mohammad-Bagher Gholivand², Zarrin Es'haghi¹

Department of Chemistry, Payam Noor University, P.O. Box 19395-4697, Tehran, Iran
Faculty of Chemistry, Razi University, Kermanshah 671496734, Iran
Received: 15 January 2022 Accepted: 12 February 2022
DOI: 10.30473/IJAC.2022.62555.1221

Abstract

For the first time, an analytical methodology based ondifferential pulse voltammetry (DPV) at a glassy carbon electrode (GCE) assisted by two multivariate calibration (MVC) models including back propagation-artificial neural network (BP-ANN), non-linear class, and partial least squares-1 (PLS-1), classical class, thatthey have been constructed on the basis of non-bilinear first order differential pulse voltammetry (DPV) data,was developed and validated for the simultaneous determination of Ascorbic acid, Uric acid, Acetaminophen, and Noradrenalinto identify which approach offers the best predictions.The baselines of the DPV signals were corrected by asymmetric least square spline regression (AsLSSR) algorithm. Before applying the PLS-1,lack of bi-linearity was tackled by potential shift correction using correlation optimised warping (COW) algorithm. The multivariate calibration (MVC) model was developed as a quaternary calibration modelin a blank human serum sample (drug-free) provided by a healthy volunteer to regard the presence of a strong matrix effect which may be caused by the possible interferents present in the serum, and it was validated and tested with two independent sets of analytes mixtures in the blank and actual human serum samples, respectively.According to the obtained results, the PLS-1 was recommended for simultaneous determination of AA, UA, AC, and NA in both blank and actual human serum samples .

Keywords

Multivariate Calibration; Ascorbic Acid; Uric aAcid; Acetaminophen; Noradrenalin; Simultaneous Determination.

1. INTRODUCTION

Ascorbic acid (AA), Uric acid (UA), Acetaminophen (AC) and Noradrenalin (NA) usually can coexist in biological matrixes and they were investigated as main molecules for biological processes in human metabolism. For example, Ascorbic acid (AA, vitamin C), a watersoluble compound, generally exists in many fruits and plants [1]. It is also can prevent free radicalinduced diseases, such as cancer and Parkinson's disease, through the reaction with free radicals [2]. Uric acid (UA) is another important biomolecule in human body, which is the primary product of purine metabolism [3].Abnormal concentration level of UA may cause such diseases as hyperuricaemia, gout and the Lesch-Nyan disease [4]. Acetaminophen (AC) is a widely used analgesic and antipyretic drug [5]. Overdoses of acetaminophen produce toxic metabolites accumulation in liver, which may cause severe and sometimes fatal hepatoxicity [6,7] and nephrotoxicity [8].Noradrenalin (NA) is a neurotransmitter belonging to a catecholamine type hormone [9] that is manufactured as a drug. Alsocalled norepinephrine, especially by those in the medical field [10].

Many methods such as high performance liquid chromatography (HPLC) [11,12], gas chromatography [13], capillary zone

*Corresponding Author: gh.mansori@gmail.com

electrophoresis [14], ion chromatography [15], biological fluids including titrimetry [16], UV-Vis spectrophotometry [17,18], have been reportedfor the determination of mentioned analyte.Nevertheless, above methodsare usually complicated time-consuming. and partly expensive. But electrochemical methods have gained much attention because of their quick response, high sensitivity, high accuracy, simple operation mode, low cost as well as ability to miniaturize. Their similar structural patterns and electrochemical signals, show overlapping when using conventional electrodes. Therefore, it is important to develop new approaches for the simultaneous determinations of these drugs. An attractive way is the use of chemometric and multivariate calibration methods. One problem which restricted the application of chemometrics in electroanalytical chemistry is the non-linearity of electrochemical data [19]. According to the literature, the shift in electrochemical responses can be originated from

adsorptive phenomena on the electrode surface, pH variations in the cell or fluctuations in the composition of cell solution [20]. However, this situation has been scarcely described for electrochemical signals. Whereaszeroth-order univariate calibration cannot detect sample components producing an interfering signal, first-

order MVC, which operates using a vector of data per sample, may compensate for these potential interferents, provided they are included in the calibration set,a property known as the "firstorder advantage" [21]. The MVC methods are increasingly used to extract relevant information from different types of absorptive spectral and electrochemical data, which have maximum information regarding the analyte of interest while discards those carrying irrelevant information (noise, saturation regions) or those heavily overlapped with other sample components which are not of interest, to predict analyte concentrations or properties of complex samples [22-24]. A basic assumption for application of linear MVC models is the data bilinearity, which may be compromised by the above commented potential shifts. Therefore several strategies have been developed for the calibration of non-linear data systems such as: a) data pretreatment or data alignment; for this reason, data alignment was performed before applying PLS-1, b)the use of linear methods (for slight nonlinearities only), c) the use of local modeling, d) the addition of extra variables, e) the use of non-linear calibration techniques [25-27] and the most important linear calibration method is PLS [28].Mathematical preprocessing techniques exist for removing variations in voltammograms from run to run, which are unrelated to analyte concentration changes [29,30]. The removal of these unwanted effects leads to more parsimonious PLS models and often produce better statistical indicators. Three type of MVC models were usually used such as: classical linear (PLS-1, CPR, and MLR), robust linear (PRM, and RCR) and non-linear (PLY-PLS, SPL-PLS, RBF-PLS, LS-SVM, WT-ANN, DWT-ANN, and BP-ANN) that we choose PLS-1 (from classical class) and BP-ANN (from non-linear class) methods for determining the concentration of the studied analytes.

In this work, we present the development of an electroanalytical methodology based on DPVat a GCE and comparison of two chemometric approachesincluding classical linear (PLS-1)and non-linear (BP-ANN)MVC models for the simultaneous determination of four studied analytesin human serum samples which have a very complex matrix.Literature survey revealed that no attempt has been made till date to the simultaneous voltammetric determination of AA, UA, AC, and NA with the aid of Chemometrics.

2. EXPERIMENTAL

2.1. Chemicals and solutions

AA, UA, AC, and NAwere purchased from Sigma-Aldrich.A phosphate buffered solution (PBS, 0.1mol L^{-1}) of pH2.0was prepared from chemicals (analytical grade) including NaH₂PO₄

and Na₂HPO₄from Merck.All other chemicals used in the investigation were of analytical grade obtained from regular sources and used without further purification.Stock standard solutions of AA, UA, AC, and NA were prepared by exact weighing and dissolving of their solid powder in distilled water (10.0 mL) with a concentration level of 0.1 mol L⁻¹and were stored at dark in a refrigerator until analysis time. Working solutions were prepared by appropriate dilution of the stock standard solutions with PBS (0.1mol L⁻¹, pH 2.0). All the solutions were prepared by doubly distilled water (DDW).

2.2. Apparatus and softwares

Electrochemical experiments were performed using anAutolab(Eco Chemie BV, Netherlands) controlled by the NOVA software (Version 1.8). A conventional three-electrode cell was used with a saturated Ag/AgCl as reference electrode, a Pt wire as counter electrode and a GCE as working electrode. The pH of the solutions was adjusted using aJENWAY-3510pH meter equipped by a combined glass electrode.All the recorded electrochemical data was smoothed, when necessary, and converted to data matrices by the use of several home-made m-files in MATLAB environment (Version 7.14, MathWorks, Inc.). All the computations for baseline correction, data alignment and multivariate calibration (MVC) were performed in MATLAB environment.All the computations were performed on a DELL XPS laptop (L502X) with Intel Core i7-2630OM 2.0GHz, 8 GB of RAM and Windows 7-64 as its operating system.

2.3. Preparation of theserum samples

A blank human serum sample (drug-free) was provided by a healthy volunteer who not exposed to any drug for at least 10.0 months. An actual human serum sample was collected from apatient under AC treatmentwhich kindly provided by a Medical Diagnostic Laboratory in Kermanshah, Iran. The following methodology was used to prepare all the serum samples: according to the method of Shu et al. [31], to eliminate protein and other substances, 5.0 mL of human serum sample was placed in a10.0 mL glass tube and 5.0 mL of 15.0 % (w/v) Zinc Sulfate solution-Acetonitrile (50/40,v/v) was added. The glass tube was vortexed for 20.0 min, maintained at 4.0 °C for 15.0 min followed by centrifugation at 4000.0 rpm for 5.0 min. Then, the supernatant was collected in the same tube and this solution was used for subsequent analyses.

2.4. Model efficiency estimation

Whether a model can be applied to analysis of human serum samples or not, model validation is

possibly the most important step in the model building sequence. In order to evaluate the performance of MVC models, each model was validated for the prediction of validation and test sets, evaluating root mean square errors of prediction (RMSEP), and relative error of prediction (REP).

$$RMSEP = \sqrt{\frac{\sum_{1}^{n} \left(y_{pred} - y_{act}\right)^{2}}{n}}$$
(1)

$$REP(\%) = \frac{100}{y_{mean}} \sqrt{\frac{1}{n} \sum_{i=1}^{n} (y_{pred} - y_{act})^2}$$
(2)

where y_{act} and y_{pred} are actual and predicted concentrations of each component, respectively, and *n* is the number of samples in validation test set.

2.5. Electrochemical procedure

Prior to electrochemical experiments, the GCE was successively polished to a mirror using 0.3 and 0.05 μ m Alumina slurry. Afterward, the electrode was washed thoroughly with ethanol and DDW and dried at room temperature. All electrochemical experiments were carried out at room temperature. The DPV measurements were carried out at the following operating conditions for the four studied analytes: step potential 0.005 V, modulation amplitude 0.025 V, modulation time 0.05 s,interval time 0.5 s, and scan rate 0.05 V s⁻¹.

3. RESULT AND DISCUSSION

3.1. Electrochemical studies

3.1.1. pH dependencestudy

To select the best pH for the simultaneous determination of AA, UA, AC, and NA, the effect of pH on the peak current of the cyclic voltammograms of AA, UA, AC, and NA was investigated. Fig. 1A-D shows the influence of the pH of the PBS (0.1mol L⁻¹), in the range of 2.0-10.0, on the signal intensities of 0.1 mM AA, UA, AC, and NA. As can be observed in Fig.1 A-D, all peak currents of the studied analytes have a maximum value at pH 2.0. Taking into account that for analytical purposes both maximal and stable currents are necessary, a pH value of 2.0 was selected for further experiments. The

oxidation peak potential of all studied analytes shifted to less positive values as the pH of the buffer solution was increased (Fig. 1A-D).



Fig. 1.Cyclic voltammograms of (A) AA $(0.1 \times 10^{-3} \text{ mol } L^{-1})$, (B) UA $(0.1 \times 10^{-3} \text{ mol } L^{-1})$, (C) AC $(0.1 \times 10^{-3} \text{ mol } L^{-1})$, and (D) NA $(0.1 \times 10^{-3} \text{mol } L^{-1})$, in 0.1mol L⁻¹ PBS at different pHs.

3.1.2. Effect of scan rate

The influences of scan rate (v) on the peak current (I_P) of AA, UA, AC, and NA at the GCE in PBS (0.1mol L⁻¹, pH 2.0) were studied by cyclic voltammetry. Fig. 2 A-D shows the influence of the scan rate (v) on the peak current (I_P) of 0.1 mM AA, UA, AC, and NA at the GCE in PBS

(0.1 mol L⁻¹, pH 2.0). In the range of 10.0-1000.0 mV s⁻¹, a linear relationship was established between $I_{\rm P}$ and $v^{1/2}$, for all of the studied analytes, indicating the diffusion controlled mechanism except UA. The linear regression equations are $I_{\rm P}$, AA(μ A)=3.7828+27.486 $v^{1/2}$ V s⁻¹(R^2 =0.9920), $I_{\rm P}$, UA(μ A)=-2.683+45.517vV s⁻¹(R^2 =0.9823), $I_{\rm P}$, AC(μ A)=0.5868-0.2653 $v^{1/2}$ V s⁻¹ (R^2 = 0.9914) and $I_{\rm P}$, NA(μ A)=-0.5157+8.8232 $v^{1/2}$ V s⁻¹(R^2 =0.9980).



Fig. 2.Cyclic voltammograms of (A) AA $(0.1 \times 10^{-3} \text{ mol } L^{-1})$, (B) UA $(0.1 \times 10^{-3} \text{ mol } L^{-1})$, (C) AC $(0.1 \times 10^{-3} \text{ mol } L^{-1})$, and (D) NA $(0.1 \times 10^{-3} \text{ mol } L^{-1})$, in 0.1 mol L^{-1} PBS (0.1 mol L^{-1} , pH 2.0) at different scan rate.Insets: dependence of $I_{\rm P}$ with scan rate.

3.2. Chemometric studies

3.2.1. A glance to necessity of MVC

Fig. 3shows the cyclic voltammograms of AA (curve a), UA (curve b), AC (curve c), NA (curve d), and their mixture (curve e) in PBS ($0.1 \text{ mol } \text{L}^{-1}$, pH 2.0). In all conditions evaluated, a strong signal overlapping was observed for the simultaneous analysis of AA, UA, AC, and NA at

the GCE (see Fig. 3, curvee). Thus, the quantification of any of these analytes will be biased if univariate calibration is used as analytical method, and for tackling this problem it was necessary to use MVC.



Fig. 3.Cyclic voltammograms of AA (curve a), UA (curve b), AC (curve c), NA (curve d), and their mixture (curve e) in PBS ($0.1 \text{ mol } \text{L}^{-1}$, pH 2.0).

3.2.2. Calibrations

3.2.2.1. Univariate calibrations

Prior to multivariate calibration experiments, calibration experiments univariate were performed (Fig. 4A-D) and calibration curves were constructed with several points as peak current versus analyte concentration in the range10.0 to 360.0×10^{-6} mol L⁻¹ for AA (inset of Fig. 4A), 5.0to 370.0×10⁻⁶ mol L⁻¹for UA (inset of Fig. 4B), 1.0 to 80.0×10-6 mol L-1 for AC (inset of Fig. 4C), 3.0 to 150.0×10⁻⁶ mol L⁻¹for NA (inset 4D), and evaluated by linear of Fig. regressionwhich were the limiting assayed concentrations in subsequent analyses. All analytes showed linear dependences between peak current and concentration at different concentrations intervals.

3.2.2.2. Multivariate calibrations

When the analytes are analyzed in the presence of interferences, the electrochemical profile revealed additional changes to those observed in the absence of interferences. The main changes observed were minor alterations in the baseline and displacement of peak potential, probably due to modification in viscosity of the solution and consequently the diffusion coefficient of the analytes. This effect produces alterations in the chemometrics responses and for this reason, the calibration and validation sets were prepared in a blank human serum sample (drug-free).

3.2.2.2.1. Calibration set

The human serum has a complex matrix and may contain a lot of unexpected interferences therefore, if the presence of these interferences was not considered during calibration, a firstorder MVC model would give biased predictions of the concentration of the analytes of interest. Therefore, the calibration set was prepared in a blank human serum sample (drug-free) which was collected from a healthy volunteer to regard the complex matrix of the serum sample which may contain a lot of unexpected interferences. This strategy was applied in order to provide PLS-1 and BP-ANNenough information concerning the signals of the analytes when they are embedded into the real background. All the calibration mixtures (the compositions of the calibration mixtures were selected according to a central composite design (CCD), (Table 1) were prepared in the blank human serum sample spiked with an appropriate amount of each analyte of interest considering the linear calibration ranges (previously established from univariate calibrations for each analyte).All samples were diluted with PBS (0.1 mol L⁻¹, pH 2.0) to adjust the pH and then appropriate amounts of these diluted samples were transferred to the electrochemical cell, and the solutions were measured in random order.Final concentration of each analyte was obtained by multiplying the detected value by the appropriate dilution factor.



Potential (V)

Fig. 4. Representative differential pulse voltammograms of (A) AA, (B) UA, (C) AC, and (D) NA,PBS (0.1 mol L^{-1} , pH 2.0)at different concentrations. Insets: dependence of I_p with concentration.

Iranian Journal of Analytical Chemistry 9 (2022) 1-14 | 6

	Table 1. Conc	entration dat	ta of calibratio	on set $(C_1 - C_{30})$,	validation s	set $(\mathbf{V_{1}}-\mathbf{V_{10}})$, and test so	et $(T_1 - T_{10})$.	
Sample	AA	UA	AC	NA	Sample	AA	UA	AC	NA
	(10-6	(10-6	(10-6	(10-6		(10-6	(10-6	(10-6	(10-6
	mol L ⁻¹)		mol L ⁻¹)	mol L ⁻¹)	mol L ⁻¹)	mol L ⁻¹)			
C_1	97.5	96.25	20.75	39.75	V_1	20	12	4	5
C_2	97.5	96.25	20.75	113.25	V 2	45	32	55	16
C_3	97.5	96.25	60.25	39.75	V 3	32	76	33	72
C_4	97.5	96.25	60.25	113.25	V_4	17	11	78	33
C_5	97.5	278.75	20.75	39.75	V 5	54	55	11	133
C_6	97.5	278.75	20.75	113.25	V 6	90	340	15	144
C ₇	97.5	278.75	60.25	39.75	V 7	250	233	22	99
C_8	97.5	278.75	60.25	113.25	V 8	345	187	76	43
C9	272.5	96.25	20.75	39.75	V 9	52	121	47	21
C_{10}	272.5	96.25	20.75	113.25	V 10	131	88	66	6
C11	272.5	96.25	60.25	39.75	T_1	78	15	32	12
C12	272.5	96.25	60.25	113.25	T_2	23	22	6	55
C ₁₃	272.5	278.75	20.75	39.75	T ₃	311	333	77	43
C_{14}	272.5	278.75	20.75	113.25	T_4	45	345	54	143
C ₁₅	272.5	278.75	60.25	39.75	T ₅	66	132	33	22
C16	272.5	278.75	60.25	113.25	T_6	22	11	11	76
C17	10	187.5	40.5	76.5	T ₇	187	17	65	54
C18	360	187.5	40.5	76.5	T_8	243	65	39	57
C19	185	5	40.5	76.5	T 9	11	54	28	88
C_{20}	185	370	40.5	76.5	T ₁₀	82	33	18	91
C ₂₁	185	187.5	1	76.5					
C22	185	187.5	80	76.5					
C ₂₃	185	187.5	40.5	3					
C ₂₄	185	187.5	40.5	150					
C25	185	187.5	40.5	76.5					
C ₂₆	185	187.5	40.5	76.5					
C ₂₇	185	187.5	40.5	76.5					
C ₂₈	185	187.5	40.5	76.5					
C29	185	187.5	40.5	76.5					
C30	185	187 5	40.5	76 5					

3.2.2.2.2.Validation set

To check the prediction ability of the model after optimizing all calibration parameters, avalidation set of ten quaternary mixtures (Table 1) was prepared in the blank human serum sample (drugfree).The concentrations of four analytes were selected at random from the corresponding calibration ranges. All samples were diluted with PBS (0.1 mol L⁻¹, pH 2.0) to adjust the pH and then appropriate amounts of these diluted samples were transferred to the electrochemical cell, and the solutions were measured in random order. Final concentration of each analyte was obtained by multiplying the detected value by the appropriate dilution factor.

3.2.2.2.3. Test set

With the purpose of evaluating the proposed method in a very interfering environment such as human serum, atest set of tenquaternary mixtures (Table 1) was prepared in the actual serum (see Section 2.3.) with random amount of each analyte of interest in the same concentration range used for calibration. All samples were diluted with

PBS (0.1 mol L⁻¹, pH 2.0) to adjust the pH and then appropriate amounts of these diluted samples were transferred to the electrochemical cell, and the solutions were measured in random order. Final concentration of each analyte was obtained by multiplying the detected value by the appropriate dilution factor.It should be noted that all the samples related to the test set were prepared in the actual serum (see Section 2.3) which contains AC therefore, the exact concentrations of the spiked AC were computed by a previous knowledge about the initial amount of AC in he serum which was obtained by analyzing the serum with HPLC-UV method prior to analyzing by the proposed method in this study.

3.2.3. Pretreatment and data arrangement

Besides the problem arising from the presence of severely overlapping analyte profiles, in the present study two additional complications may occur: (1) interactions among analytes and the background interferents present in the serum, which may cause signal changes in comparison with pure analyte profiles, and (2) sample-to-

sample potential shifts in the analyte profiles, which are common in voltammetric studies.For tackling the first problem, it was necessary to include the possible interferents in the calibration set in order to allow a first order algorithm to model the analyte-background interactions before prediction on new samples. Concerning the second commented problem, some preprocessing alternatives were independently applied on the electrochemical responses before model building and validation [21,32,33,34].

In our previous work [21] we pointed out that voltammetric performance can be enhanced by eliminating noise and background components therefore, baseline elimination is a crucial step for reducing both complexity and number of the unexpected components. Moreover, it was demonstrated that the use of signal pre-treatments such as baseline- and potential shift-corrections improve the quality of first-order voltammetric signals and, as a consequence, the performance of resolution by first-order algorithms [21]. In the next sections, some strategies areexaminedto achieving the mentioned aims.

3.2.3.1. Baseline correction

Baseline correction has been considered as a critical step for enhancing the signals and reducing the complexity of the analytical data [35]. Considering this aim, we used the method proposed by Eilers et al. [36] for background elimination in two-dimensional signals based on asymmetric least squares splines regression approach. In the asymmetric least squares method [37] the following cost function is minimized: $Q = \sum_{i} \upsilon_i (y_i - f_i)^2 + \lambda \sum_i (\Delta^2 f_i)^2$

where *y* is the experimental signal, f is a smooth approximation of baseline trend (y), Δ is the derivative of f, i denotes successive values of the signal, the positive parameter λ is a regularization parameter that weigh the second term and vare weights. The positive deviations from the estimated baseline (peaks) have lowv values while the negative deviations (baseline) obtain highv values. In the multidimensional extension of baseline correction method, Eilers et al. proposed the splines-based approach to smoothing instrumental signal (the penalty term in Eq. 1). Details of the implementation of the mentioned method can be found in the literature [36].

3.2.3.2. Potential shift correction

For chemometric model building, several strategies have been proposed to align shifted signals such as chromatograms, electropherograms or NIR spectra. One of the most popular ones is correlation optimised warping (COW) [38,39]. However, this situation

has been scarcely described for electrochemical signals [21]. According to the literature, the shift in electrochemical responses can be originated from adsorptive phenomena on the electrode surface, pH variations in the cell or fluctuations in the composition of cell solution, among others [20].

A basic assumption for application of a multivariate calibration model is the data bilinearity, which may be compromised by the above commented potential shifts. Therefore, the DPV signals were aligned towards a target signal using COW. The COW algorithm was introduced by Nielsen et al. [38] as a method to correct for shifts in discrete data signals. It is a piecewise or segmented data preprocessing technique that uses dynamic programming to align a sample signal towards a reference signal by stretching or compression of sample segments using linear interpolation. First, the segment and slack were optimized using a simplex-like optimization routine and then mean voltammogram was selected as target"signal".

The results of baseline and shift corrections are shown in Fig. 5. Fig. 5A-C shows the raw DPV data recorded for the calibration, validation and test sets, respectively. Fig. 5D-Fshow the results baseline-corrected data of calibration, of validation and test sets, respectively, and as can be seen the baselinesare satisfactorily corrected. Fig. 5G-I show the results of applying COW for data alignment and it confirms the capability of COW for aligning the data. The baseline- and shift-corrected data was used to model building by PLS-1 while the baseline-corrected data without shift-correction was used to model building by BP-ANN.

3.3. Performance evaluation of PLS-1 and BP-ANN in blank and actual serum samples

In MVC calibration, it is usual to have two data sets: a calibration set, employed to build the regression model, and a validation set to check the prediction ability of the model after all calibration parameters have been optimized, but in this study we are going to determine the analytes' concentrations in actual human serum samples which have a very complex matrix due to the presence of many interferents present in the serum therefore, with the purpose of evaluating the method in the presence of interferents, a test set was also prepared in anactual human serum sample.

Before calibration with PLS-1, it is usual to assess the optimum number of latent variables in order to avoid overfitting, by applying the well-known cross-validation method described by Haaland and Thomas [40]. With the purpose of estimating the number of optimum latent variables for ACO-GA-SS-PLS-1 leave-one-sample-out crossvalidation was performed.



B C A Current (µAL. Current (µ.A. Current LAA 03 0.4 0.1 02 05 05 0.7 0.8 02 03 04 05 0.5 U 15 0.4 05 0.5 07 08 11 02 0.3 Potential M Potential (N) Potential (M) 25 D E F Urrent (µAL Current (MA) urrent (µA 01 81 02 0.3 0.4 05 0.5 0.7 0.8 0 21 02 0.3 0.4 05 16 17 08 ٥ 01 02 0.3 0.4 0.5 0.5 0.7 08 0 Potential (V) Potential (V) Potential (M) 3 25 G I Н Currens (µA). Current (MA) urrent LA 0.4 0.5 0.6 0.7 0.8 05 05 0.7 0.8 0.1 02 03 01 02 03 0.4 0 0.1 02 03 04 05 06 07 08 Potential M Potential (V) Potential (M)

Iranian Journal of Analytical Chemistry 9 (2022) 1-14 | 8

Fig. 5. (A)-(C): Raw data related to calibration, validation, and test sets, respectively, (D)-(F):Baseline corrected data by AsLSSR related to calibration, validation, and test sets, respectively, and (G)-(I): shift corrected data by COW related to calibration, validation, and test sets, respectively.

The optimum number of factors is estimated by calculating the ratios F (A) = PRESS (A<A*)/PRESS (A), where PRESS = Σ ($c_{i,acr}$ - $c_{i,pred}$)², A is a trial number of factors, A* corresponds to the minimum PRESS, and $c_{i,acr}$ and $c_{i,pred}$ are the actual and predicted concentrations for the *i*th sample left out of the calibration during cross validation, respectively. Then, the number of factors leading to a probability of less than 75% that F > 1 is selected. This analysis led to the conclusion that the latter number is 4 for all the cases, as expected for this system using a mean

centering procedure. The presence of nonlinearities which are usually found in electrochemical data causes more latent variables to model the variability of the data but in the present study this limitation was tackled by potential shift correction.Then, the PLS-1 algorithm was run on the calibration, and validationsets.According to the results presented in Table 2, with very satisfactory values for RMSEPs and REPs for the four analytes of interest it is apparent that the PLS-1 approach has found the correct answer.

	PLS-1				BP-ANN	1		
Validation set	AA	UA	AC	NA	AA	UA	AC	NA
V_1	19.8	11.9	4.03	5.02	18.85	12.98	4.5	5.4
V_2	45.3	32	55.9	15.99	47.1	34.6	57.3	17.1
V ₃	31.8	76	33.6	72.03	34.3	73.2	31.2	74.9
V_4	16.92	11.01	79.2	33.1	18.9	10	76.5	35.6
V5	54.9	55.1	11.3	133	55.4	52.1	10.2	137.9
V_6	90.1	340.1	14.9	144	92.3	349.9	14.1	152.6
\mathbf{V}_7	250.3	234.1	21.6	99.1	252.3	238.7	20.1	104.5
V_8	345.4	188.1	76.5	43.01	346.7	190.9	73.2	47.8
V9	52.6	120.1	47.9	21.01	54.9	126.4	49.6	24.3
V ₁₀	132.03	87.5	65.8	6.02	138.5	90.8	68.9	5.02
RMSEP ()	0.5164	0.5924	0.6301	0.0469	3.0777	4.5571	1.9773	4.2385
REP ()	0.4985	0.5129	1.5482	0.0820	2.9707	3.9456	4.8584	7.4099
	PI S_1					т		
	1 LS-1				BP-ANN	N		
Test set	AA	UA	AC	NA	AA	UA	AC	NA
Test set T ₁	AA 78.2	UA 15.2	AC 32.4	NA 11.9	AA 80.1	UA 16.3	AC 30.1	NA 10.2
Test set T ₁ T ₂	AA 78.2 23.2	UA 15.2 22.3	AC 32.4 5.9	NA 11.9 55.07	AA 80.1 21.1	UA 16.3 24.3	AC 30.1 7.1	NA 10.2 58.8
Test set T1 T2 T3	AA 78.2 23.2 311.8	UA 15.2 22.3 333.6	AC 32.4 5.9 77.3	NA 11.9 55.07 42.9	BP-ANN AA 80.1 21.1 307.3	UA 16.3 24.3 339.8	AC 30.1 7.1 75.2	NA 10.2 58.8 46.3
Test set T ₁ T ₂ T ₃ T ₄	AA 78.2 23.2 311.8 44.2	UA 15.2 22.3 333.6 345.3	AC 32.4 5.9 77.3 53.5	NA 11.9 55.07 42.9 142.9	BP-ANN AA 80.1 21.1 307.3 43.1	UA 16.3 24.3 339.8 349.9	AC 30.1 7.1 75.2 42.3	NA 10.2 58.8 46.3 140.1
Test set T1 T2 T3 T4 T5	AA 78.2 23.2 311.8 44.2 65.1	UA 15.2 22.3 333.6 345.3 132.1	AC 32.4 5.9 77.3 53.5 32.6	NA 11.9 55.07 42.9 142.9 21.9	BP-ANN AA 80.1 21.1 307.3 43.1 64.2	UA 16.3 24.3 339.8 349.9 138.9	AC 30.1 7.1 75.2 42.3 31.2	NA 10.2 58.8 46.3 140.1 19.3
Test set T1 T2 T3 T4 T5 T6	AA 78.2 23.2 311.8 44.2 65.1 21.7	UA 15.2 22.3 333.6 345.3 132.1 11.2	AC 32.4 5.9 77.3 53.5 32.6 10.8	NA 11.9 55.07 42.9 142.9 21.9 75.9	BP-ANN AA 80.1 21.1 307.3 43.1 64.2 20.1	UA 16.3 24.3 339.8 349.9 138.9 13.2	AC 30.1 7.1 75.2 42.3 31.2 9.9	NA 10.2 58.8 46.3 140.1 19.3 79.9
Test set T1 T2 T3 T4 T5 T6 T7	AA 78.2 23.2 311.8 44.2 65.1 21.7 188.5	UA 15.2 22.3 333.6 345.3 132.1 11.2 17.2	AC 32.4 5.9 77.3 53.5 32.6 10.8 64.3	NA 11.9 55.07 42.9 142.9 21.9 75.9 53.2	BP-ANN AA 80.1 21.1 307.3 43.1 64.2 20.1 190.8	UA 16.3 24.3 339.8 349.9 138.9 13.2 19.4	AC 30.1 7.1 75.2 42.3 31.2 9.9 63.2	NA 10.2 58.8 46.3 140.1 19.3 79.9 51.1
Test set T1 T2 T3 T4 T5 T6 T7 T8	AA 78.2 23.2 311.8 44.2 65.1 21.7 188.5 244.1	UA 15.2 22.3 333.6 345.3 132.1 11.2 17.2 65.6	AC 32.4 5.9 77.3 53.5 32.6 10.8 64.3 38.5	NA 11.9 55.07 42.9 142.9 21.9 75.9 53.2 56.4	BP-ANN AA 80.1 21.1 307.3 43.1 64.2 20.1 190.8 249.6	UA 16.3 24.3 339.8 349.9 138.9 13.2 19.4 62.3	AC 30.1 7.1 75.2 42.3 31.2 9.9 63.2 37.8	NA 10.2 58.8 46.3 140.1 19.3 79.9 51.1 53.2
Test set T1 T2 T3 T4 T5 T6 T7 T8 T9	AA 78.2 23.2 311.8 44.2 65.1 21.7 188.5 244.1 11.4	UA 15.2 22.3 333.6 345.3 132.1 11.2 17.2 65.6 53.2	AC 32.4 5.9 77.3 53.5 32.6 10.8 64.3 38.5 27.8	NA 11.9 55.07 42.9 142.9 21.9 75.9 53.2 56.4 87.8	BP-ANN AA 80.1 21.1 307.3 43.1 64.2 20.1 190.8 249.6 12.1	UA 16.3 24.3 339.8 349.9 138.9 13.2 19.4 62.3 52.1	AC 30.1 7.1 75.2 42.3 31.2 9.9 63.2 37.8 26.2	NA 10.2 58.8 46.3 140.1 19.3 79.9 51.1 53.2 91.2
Test set T1 T2 T3 T4 T5 T6 T7 T8 T9 T10	AA 78.2 23.2 311.8 44.2 65.1 21.7 188.5 244.1 11.4 82.4	UA 15.2 22.3 333.6 345.3 132.1 11.2 17.2 65.6 53.2 32.4	AC 32.4 5.9 77.3 53.5 32.6 10.8 64.3 38.5 27.8 17.66	NA 11.9 55.07 42.9 142.9 21.9 75.9 53.2 56.4 87.8 90.8	BP-ANN AA 80.1 21.1 307.3 43.1 64.2 20.1 190.8 249.6 12.1 85.9	UA 16.3 24.3 339.8 349.9 138.9 13.2 19.4 62.3 52.1 31.1	AC 30.1 7.1 75.2 42.3 31.2 9.9 63.2 37.8 26.2 16.9	NA 10.2 58.8 46.3 140.1 19.3 79.9 51.1 53.2 91.2 88.8
Test set T1 T2 T3 T4 T5 T6 T7 T8 T9 T10 RMSEP ()	AA 78.2 23.2 311.8 44.2 65.1 21.7 188.5 244.1 11.4 82.4 0.7771	UA 15.2 22.3 333.6 345.3 132.1 11.2 17.2 65.6 53.2 32.4 0.4505	AC 32.4 5.9 77.3 53.5 32.6 10.8 64.3 38.5 27.8 17.66 0.4007	NA 11.9 55.07 42.9 142.9 21.9 75.9 53.2 56.4 87.8 90.8 0.3368	BP-ANN AA 80.1 21.1 307.3 43.1 64.2 20.1 190.8 249.6 12.1 85.9 3.2648	UA 16.3 24.3 339.8 349.9 138.9 13.2 19.4 62.3 52.1 31.1 3.8723	AC 30.1 7.1 75.2 42.3 31.2 9.9 63.2 37.8 26.2 16.9 3.9815	NA 10.2 58.8 46.3 140.1 19.3 79.9 51.1 53.2 91.2 88.8 3.1210

Gh. Mansori et al./ Application of Non-bilinear Voltammetric Data | 9

Table 2. Results of application of PLS-1 and BP-ANN to validation and test sets.

Table 3. ANN training results for all data sets.

AA	Value
Architecture (input-hidden-output neurons)	2000-05-01
Number of training epoch	450-1200
Learning rate	0.5
Momentum	0.5
UA	Value
Architecture (input-hidden-output neurons)	2005-04-01
Number of training epoch	400-8735
Learning rate	0.5
Momentum	0.5
AC	Value
Architecture (input-hidden-output neurons)	2000-06-01
Number of training epoch	450-10080
Learning rate	0.5
Momentum	0.5
NA	Value
Architecture (input-hidden-output neurons)	2004-06-01
Number of training epoch	600-9765
Learning rate	0.5
Momentum	0.5

With the purpose of analyzing the potentiality of the evaluated methodology based on DPV data processed by PLS-1, a test set of ten quaternary mixtures (see Table 1) was prepared in an actual serum (see Section 3.2.2.2.3.).According to the results presented in Table 2 for the test set, with very satisfactory values for RMSEP and REP for the four analytes of interest it is apparent that the PLS-1 approach has found the correct answer.

Intrinsically non-linear data can be processed with non-linear methods. Therefore BP-ANN was also used to predict the validation and test sets using the training parameters reported in Table 3. According to the results presented in Table 2 for both validation and test sets, satisfactory values for RMSEP and REP for the four analytes of interest were not founded and it shows that the BP-ANN approach has not found the correct answer.

Therefore, in this study PLS-1 shows better results than BP-ANN therefore, for the sake of a further investigation into the accuracy of the method (PLS-1), proposed thepredicted concentrations of both validation and test sets were regressed on the nominal concentrations. In this case an ordinary least squares (OLS) analysis of predicted concentrations versus nominal concentrations was applied [41]. The calculated intercept and slope were compared with their theoretically expected values (intercept = 0, slope = 1), based on the elliptical joint confidence region (EJCR) test. If the ellipses contain the values 0 and 1 for intercept and slope (ideal point), respectively, showing the predicted and nominal values do not present significant difference at the level of 95% confidence and the elliptic size denotes precision of the analytical method, smaller size corresponds to higher precision [42]. Fig. 6A-Dand Fig. 6F-I show the regression of predicted concentrations on nominal values based on OLS method corresponding to the validation and test sets, respectively, and Fig. 6E and Fig. 6J show the corresponding ellipses of the EJCR analyses for validation and test sets, respectively.As can be concluded from Fig. 6A-D and Fig. 6F-I, the predictions for AA, UA, AC, and NA in both validation and test sets are in good agreement with the nominal values. If the EJCRs for validation set are analyzed (Fig. 6E), it is notable that while the ellipses for AA, UA, and NA include the theoretically expected point (ideal point), indicating accuracy of the developed methodology for these analytes, the ideal point falls on the AC ellipse (red ellipse), denoting slightly poorer prediction accuracy for AC. But all the ellipses in the test set contain the ideal point (Fig. 6J) which shows the accurate determination of all analytes in test set by the developed methodology. The statistical results shown in Table 2, with adequate values for RMSEPand REP for all analytes, do also support this conclusion.



Fig. 6. (A)-(D) and (F)-(I): plots for predicted concentrations as a function of nominal values for AA, UA, AC, and NAin validation and test set, respectively. (E) and (J):elliptical joint regions (at 95% confidence level) for the slopes and intercepts of the regressions for AA (blue ellipse), UA (yellow ellipse), AC (red ellipse), and NA (green ellipse) in validation and test set, respectively. Plus point(+) marks the theoretical (0,1) point.

4. CONCLUSION

In the present article, combination of electrochemistry with chemometrics led us to introduce an efficient analytical method for simultaneous determination of AA, UA, AC, and NA at a GCE in very interfering media. The four studied analytes exhibited a strong voltammetric overlapping for the simultaneous analysis of these compounds. The overlapping was more successfully resolved using PLS-1 with preprocessing than BP-ANN. The baseline of the DPV signals was successfully removed by an efficient chemometric algorithm. Because of the non-bilinear behavior of the experimental data, the potential shift correction was carried out by COW as an efficient chemometric algorithm before applying PLS-1.To regard the presence of a strong matrix effect which may be caused by the possible interferents present in the human serum sample, the MVC models were built and validated in a blank human serum sample (drug-free) provided by a healthy volunteer which allowed us to exploiting first-order advantage for the simultaneous determination of the four studied analytes in very interfering media such as human serum samples. Finally, the application of the developed method based on PLS-1 to simultaneously assay the concentrations of AA, UA, AC, and NA in an actual human serum sampleallowed to obtain satisfactory results. This study allows one to propose the present method as a promissory, cheap and accessible alternative for routine determination of the concentrations of AA, UA, AC, and NA in human serum samples.

Acknowledgements

The financial support of this project by Razi University Research Council is gratefully acknowledged.

REFERENCES

- Y. Andreu, S.D. Marcos, J.R. Castillo and J. Galbán, Sensor film for Vitamin C determination based on absorption properties of polyaniline, *Talanta* 65 (2005) 1045-1051.
- [2] P.J. O'Connell, C. Gormally, M. Pravda and G.G. Guilbault,Development of an amperometric 1-ascorbic acid (Vitamin C) sensor based on electropolymerised aniline for pharmaceutical and food analysis, *Anal. Chim. Acta*, 431 (2001) 239-247.
- [3] X.X. Weng, Q.X. Cao, L.X. Liang, J.G. Chen, C.P. You, Y.M. Ruan, H.J. Lin and L.J.Wu, Simultaneous determination of dopamine and uri acid using layer-by-layer graphene and chitosan assembled multilayer films, *Talanta*, 117 (2013) 359-365.

- [4] D.V.V.S. Eswara and H.A. Mottola, Determination of Uric Acid at the Microgram Level by a Kinetic Procedure Based on a "Pseudo-Induction" Period, *Anal. Chem.*, 46 (1974) 1777-1781.
- [5] J.G.M. Bessems and N.P.E. Vermeulen, Paracetamol (Acetaminophen)-Induced Toxicity: Molecular and Biochemical Mechanisms, Analogues and Protective Approaches, *Crit. Rev. Toxicol.* 31 (2001) 55-138.
- [6] M.T. Olaleye and B.T.J. Rocha, Acetaminophen-induced liver damage in mice: Effects of some medicinal plants on the oxidative defense system, *Exp. Toxicol. Pathol.* 59 (2008) 319-327.
- [7] J.C. Roberts, H.L. Phaneuf, J.G. Szakacs, R.T. Zera, J.G. Lamb and M.R. Franklin, Differential Chemoprotection against Acetaminophen-Induced Hepatotoxicity by Latentiated L-Cysteines, *Chem. Res. Toxicol.* 11 (1998) 1274-1282.
- [8] M. Mazer and J. Perrone, Acetaminophen-Induced Nephrotoxicity: Pathophysiology, Clinical Manifestations, and Management, J. *Med. Toxicol.* 4 (2008) 2-6.
- [9] H. Beitollahi, H. Karimi-Maleh and H. Khabazzadeh, Nanomolar and Selective Determination of Epinephrine in the Presence of Norepinephrine Using Carbon Paste Electrode Modified with Carbon Nanotubes and Novel 2-(4-Oxo-3-phenyl-3,4-dihydroquinazolinyl)-N'-phenyl-hydrazinecarbothioamide, Anal. Chem. 80 (2008) 9848-9851.
- [10] http://www.wisegeek.org/what-isnoradrenaline.htm; 26, May, 2013.
- [11] M.G. Gioia, P. Andreatta, S. Boschetti and R. Gatti, Development and validation of a liquid chromatographic method for the determination of ascorbic acid, dehydroascorbic acid and acetaminophen in pharmaceuticals, J. Pharm. Biomed. Anal. 48 (2008) 331-339.
- [12] V. Carrera, E. Sabater, E. Vilanova and M.A. Sogorb, A simple and rapid HPLC–MS method for the simultaneous determination of epinephrine, norepinephrine, dopamine and 5-hydroxytryptamine: Application to the secretion of bovine chromaffin cell cultures, *J. Chromatogr. B*, 847 (2007) 88-94.
- [13] D.L. Kuhlenbeck, T.P.O. Neill, C.E. Mack, S.H. Hoke and K.R. Wehmeyer, Determination of norepinephrine in small volume plasma samples by stable-isotope dilution gas chromatography-tandem mass spectrometry with negative ion chemical

ionization, J. Chromatogr. B, 738 (2000) 319-330.

- [14] T. Wu, Y. Guan and J. Ye, Determination of flavonoids and ascorbic acid in grapefruit peel and juice by capillary electrophoresis with electrochemical detection, *Food Chem.*, 100 (2007) 1573-1579.
- [15] C.L. Guan, J. Quyang, Q.L. Li, B.H. Liu and W.R.G. Baeyens, Simultaneous determination of catecholamines by ion chromatography with direct conductivity detection, *Talanta*, 50 (2000) 1197-1203.
- [16] K.G. Kumar and R. Letha, Determination of Paracetamol in pure form and in dosage forms using *N*,*N*-dibromo dimethylhydantoin, *J. Pharm. Biomed. Anal.*, 15 (1997) 1725-1728.
- [17] V. Rodenas, M.S. Garcia, C. Sanchez-Pedreno and M.I. Albero, Simultaneous determination of propacetamol and paracetamol by derivative spectrophotometry, Talanta, 52 (2000) 517-523.
- [18] N. Erk, Y. Ozkan, E. Banoglu, S.A. Ozkan and Z. Senturk, Simultaneous determination of paracetamol and methocarbamol in tablets by ratio spectra derivative spectrophotometry and LC, J. Pharm. Biomed. Anal., 24 (2001) 469-475.
- [19] A. Alberich, J.M. Diaz-Cruz, C. Arino and M. Esteban, Potential shift correction in multivariate curve resolution of voltammetric data. General formulation and application to some experimental systems, Analyst, 133 (2008) 112-125.
- [20] A.J. Bard, L.R. Faulkner, ELECTROCHEMICAL METHODS Fundamentals *and Applications*, John Wiley & Sons, Inc., New York, 2001.
- [21] M.B. Gholivand, A.R. Jalalvand and H.C. Goicoechea, Th. Skov, Chemometricsassisted simultaneous voltammetric determination of ascorbic acid, uric acid, dopamine and nitrite: Application of nonbilinear voltammetric data for exploiting first-order advantage, *Talanta*, 119 (2014) 553-563.
- [22] L. Xu, J.H. Jiang, H.L. Wu, G.L. Shen and R.Q. Yu, Variable-weighted PLS, *Chem. Intell. Lab. Syst.*, 85 (2007) 140-143.
- [23] S. Wold, H. Martens and H. Wold, The multivariate calibration problem in chemistry solved by the PLS method, Springer-Verlag, Heidelberg, 1983.
- [24] S. Wold, J. Cheney, N. Kettaneh and C. McCready, The chemometric analysis of point and dynamic data in pharmaceutical and biotech production (PAT) — some

objectives and approaches, *Chem. Intell. Lab. Syst.*, 84 (2006) 159-163.

- [25] H. Yang, P.R. Griffiths and J.D. Tate, Comparison of partial least squares regression and multi-layerneural networks for quantification of nonlinear systems and application to gas phase Fourier transform infrared spectra, *Anal. Chim. Acta*, 489 (2003) 125-136.
- [26] V. Centner, J. Verdú-Andrés, B. Walczak, D. Jouan-Rimbaud, F. Despagne, L. Pasti, D.L. Massart and O.E. de Noord, Comparison of Multivariate Calibration Techniques Applied to Experimental NIR Data Sets, *Appl. Spectrosc.*, 54 (2000) 608-623.
- [27] T. Naes, T. Isakson, T. Fearn, T. Davies, A User-Friendly Guide to Multivariate Calibration and Classification, NIR. Publications, Chichester, 2002.
- [28] S.Wold, M. Sjostrom and L. Eriksson, PLSregression: a basic tool of chemometrics, *Chemom. Intell. Lab. Syst.*, 58 (2001) 109-130.
- [29] P. Geladi, D. MacDougall and H. Martens, Linearization and Scatter-Correction for Near-Infrared Reflectance Spectra of Meat, Appl. Spectrosc., 39 (1985) 491-500.
- [30] R.J. Barnes, M.S. Dhanoa and S.J. Lister, Standard Normal Variate Transformation and De-trending ofNear-Infrared Diffuse Reflectance Spectra, Appl. Spectrosc., 43 (1989) 772-777.
- [31] J.H. Suh, Y.Y. Lee, H.J. Lee, M. Kang, Y. Hur, S.N. Lee, D.H. Yang and S.B. Han, Dispersive liquid–liquid microextraction based on solidification of floating organic droplets followed by high performance liquid chromatography for the determination of duloxetine in human plasma, J. Pharm. Biomed. Anal., 75 (2013) 214-219.
- [32] M.B. Gholivand, A.R. Jalalvand, H.C. Goicoechea and R. Gargallo, Th. Skov, G. Paimard, Combination of electrochemistry with chemometrics to introduce an efficient analytical method for simultaneous quantification of five opium alkaloids in complex matrices, Talanta, 131 (2015) 26-37.
- [33] M.B. Gholivand, A.R. Jalalvand and H.C. Goicoechea, Th. Skov, Generation of nonmultilinear three-way voltammetric arrays by an electrochemically oxidized glassy carbon electrode as an efficient electronic device to achieving second-order advantage: Challenges, and tailored applications, *Talanta*, 134 (2015) 607-618.
- [34] M.B. Gholivand, A.R. Jalalvand and H.C. Goicoechea, Th. Skov, Advanced and

tailored applications of an efficient electrochemical approach assisted by AsLSSR–COW–rPLS and finding ways to cope with challenges arising from the nature of voltammetric data, *Chemom. Intell. Lab. Syst.*, 146 (2015) 437-446.

- [35] M.M. De Zan, M.D. Gil García, M.J. Culzoni, R.G. Siano, H.C. Goicoechea and M. Martínez Galera, Solving matrix-effects exploiting the second order advantage in the resolution and determination of eight tetracycline antibiotics in effluent wastewater by modelling liquid chromatography data with multivariate curve resolution-alternating least squares and unfolded-partial least squares followed by residual bilinearization algorithms I. Effect of signal pre-treatment, J. Chromatogr. A, 1179 (2008) 106-114.
- [36] P.H.C. Eilers, I.D. Currie and M. Durban, Fast and compact smoothing on large multidimensional grids, *Comput. Statist. Data Anal.*, 50 (2006) 61-67.
- [37] P.H.C. Eilers, Parametric Time Warping, Anal. Chem. 76 (2004) 404-411.
- [38] N.P.V. Nielsen, J.M. Carstensen and J. Smedsgaard, Aligning of single and multiple wavelength chromatographic profiles for chemometric data analysis using correlation optimized warping, J. Chromatogr. A, 805 (1998) 17-35.
- [39] G. Tomasi, F. Van den Berg and C. Andersson, Correlation optimized warping and dynamic time warping as preprocessing methods for chromatographic data, *J. Chemom.*, 18 (2004) 231-241.
- [40] D.M. Haaland and E.V. Thomas, Partial Least-Squares Methods for Spectral Analyses. 1. Relation to Other Quantitative Calibration Methods and the Extraction of Qualitative Information, *Anal. Chem.*, 60 (1988) 1193-1202.
- [41] A.G. Gonzalez, M.A. Herrador and A.G. Asuero, Intra-laboratory testing of method accuracy from recovery assays, *Talanta*, 48 (1999) 729-736.
- [42] J.A. Arancibia, G.M. Escandar, Two different strategies for the fluorimetric determination of piroxicam in serum, *Talanta*, 60 (2003) 1113-1121.



COPYRIGHTS

© 2022 by the authors. Lisensee PNU, Tehran, Iran. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution 4.0 International (CC BY4.0) (http://creativecommons.org/licenses/by/4.0)



Iranian Journal of Analytical Chemistry 9 (2022) 1-14 | 14

کاربرد داده های ولتامتری غیر دو خطی به دلیل برتری روش های مرتبه اول در اندازه گیری همزمان آسکوربیک اسید، اوریک اسید، استامینوفن و نورآدرنالین در محیط تداخل

قباد منصوری *، محمدباقر قلیوند ، زرین اسحاقی ا

۲. گروه شیمی، دانشگاه پیام نور، صندوق پستی ۴۶۹۷–۱۹۳۹، تهران، ایران
۲. دانشکده شیمی، دانشگاه رازی، کرمانشاه، ایران
تاریخ دریافت: ۲۵ دی ۱٤۰۰ تاریخ پذیرش: ۲۳ بهمن ۱٤۰۰

چکیدہ

یک روش تجزیه ای بر اساس ولتامتری پالس تفاضلی در سطح الکترود کربن شیشه ای با همکاری دو روش چند متغیره کمومتریکسی شامل پس انتشار – شبکه عصبی مصنوعی که از روشهای غیر خطی می باشد و روش حداقل مربعات جزئی ۱ – که از روشهای کلاسیک است و در هر دو روش از داده های ولتامتری پالس تفاضلی که قاعدتاً دادهای مرتبه اول غیر خطی هستند استفاده گردیده است، برای اندازه گیری همزمان اسکوربیک اسید، اوریک اسید، استامینوفن و نورآدرنالین توسعه و اعتبار سنجی می شوند تا مشخص گردد که کدام روش پیش بینی بهتری را برای غلظت گونه های یاد شده ارائه می دهد. خط زمینه طیفهای ولتامتری پالس تفاضلی توسط الگوریتم رگرسیون کوچکترین مربعات نامتقارن تصحیح گردید. جابجایی پتانسیل و مشکل دوخطی بودن دادهها، پیش از استفاده از روش حداقل مربعات جزئی ۱ – با به کارگیری الگوریتم کوچکترین مربعات نامتقارن تصحیح گردید. جابجایی پتانسیل و مشکل دوخطی بودن دادهها، پیش از استفاده از روش حداقل مربعات جزئی ۱ – با به کارگیری الگوریتم کوچکترین مربعات نامتقارن تصحیح گردید. جابجایی پتانسیل و مشکل دوخطی بودن دادهها، پیش از استفاده از روش حداقل مربعات جزئی ۱ – با به کارگیری الگوریتم مستگی انحراف بهینه شده اصلاح شد. مدل کالیبراسیون چند متغیره به عنوان یک مدل کالیبراسیون چهارتایی در یک نمونه سرم انسانی خالی (بدون دارو) ارائه شده توسط یک داوطلب سالم برای در نظر گرفتن وجود یک اثر ماتریکس قوی که ممکن است توسط تداخلهای احتمالی موجود در سرم ایجاد شود، ایجاد شد و با دو مجموعه مستقل از مخلوط آنالیتها در نمونههای خالی و واقعی سرم انسانی اعتبارسنجی و آزمایش شد. در نهایت، روش حداقل مربعات جزئی ۱ – برای اندازه گیری همزمان اسکوربیک اسید، اوریک اسید، استامینون و نورآدرنالین هم در نمونه سرم انسانی مورد بررسی نتایج بهتری را در مقایسه با پس انتشار – شبکه عصبی مصنوعی نشان داد.

واژههای کلیدی

روش چند متغیره؛ اسکوربیک اسید، اوریک اسید؛ استامینوفن و نور آدرنالین؛ اندازه گیری همزمان.