

## GPU-Based Adaptive Beamformer for Medical Ultrasound Imaging

Yaser Fathi<sup>a</sup>, Ali Mahloojifar<sup>a</sup>, Babak Mohammadzadeh Asl<sup>a</sup>

<sup>a</sup> Tarbiat Modares University, Iran

Minimum Variance (MV) adaptive beamforming technique is attractive as it is capable of producing high quality images with narrow mainlobe width and low sidelobe level. However, because of its substantially higher computational requirements, realizing MV in real-time is difficult. Recent advancements in commodity Graphical Processing Units have made very high performance computing possible at very affordable price. The use of GPUs, which are programmable devices with a high level of parallelism, can accelerate the computations of the beamforming process. In this paper, our primary results of implementing simple adaptive bemformer on GPU were presented.

66 ۳۰ آذرو ۱ دی ۱۳۹۱ - دانشگاه صنعتی امیرکبیر (بلی تکنیک تهران)

December 20-21, 2012, Amirkabir University of Technology (Tehran Polytechnic), Tehran, Iran. http://www.icbme.ir